

SOUTHWEST RESEARCH INSTITUTE
Department of Quality Assurance
Calibration Laboratory • 522-5215

WORK ORDER

WORK ORDER # 36232 ASSET # 007580 DATE 30 Sept 99

ITEM DATA:

Manufacturer S.M. Gauge Model 0-60 mbar / 0117-9654615
Description pressure gauge Serial # 1580
Accessories _____

ACTION REQUESTED cal / cal before use

CUSTODIAN DIV 30, Anna Maria

Turned in by: Mark Thurston Phone 3530

CHARGE # 20 01402 159 Date Required _____

INSTRUMENT USED ON: (DOD/NASA) (NUCLEAR) (GLP) (SPPE) (ISO)
 OTHER _____

COPY OF CALIBRATION CERTIFICATE (Yes) (No)

NEW WORK Yes No If yes, an evaluation shall be made to verify capabilities.

By MAW Date 09-30-99

Work involves proprietary/confidential information or equipment (Yes) (No)

CONDITION RECEIVED: _____ Out of tolerance
_____ In tolerance
 Damaged (Contact customer)
_____ Contact _____ Date _____
_____ Received into system, introduced or reactivated

ACTION TAKEN: (Calibration/Repair/Parts) _____

* Gauge Rejected. Sticks badly at about 55 mbar.
Could be dangerous to project material!

* Replacement Suggested

CAL ENVIRONMENT: Temperature _____ °F Humidity _____ %RH

CALIBRATED/REPAIRED:
By MAW Cal Procedure _____
Date 10-6-99 Accuracy _____
Cal Interval _____ Reliability Code _____
Next Cal Due _____ Cal Time 0.5 hr Repair Time _____
Standards used (Asset #) _____

DATE COMPLETED 10-6-99

DATE PICKED UP 10-7-99 PICKED UP BY Marie LESEUNE

36232

