

SOUTHWEST RESEARCH INSTITUTE

Department of Quality Assurance

Calibration Laboratory • 522-5215

WORK ORDER

WORK ORDER # 36231 ASSET # 007581 DATE 30 Sept 99

ITEM DATA:

Manufacturer S.M. Gauge Model 0-100 mbar / 0117-9654615

Description vacuum pressure gauge Serial # 7581

Accessories _____

ACTION REQUESTED cal / cal before use

CUSTODIAN Div. 20, Anne Marie

Turned in by: Mark Thurston Phone 3530

CHARGE # 20-01402.159 Date Required _____

INSTRUMENT USED ON: (DOD/NASA) (NUCLEAR) (GLP) (SPPE) (ISO)
 OTHER _____

COPY OF CALIBRATION CERTIFICATE (Yes) (No)

NEW WORK Yes No If yes, an evaluation shall be made to verify capabilities.

By MMS Date 09-30-99

Work involves proprietary/confidential information or equipment (Yes) (No)

CONDITION RECEIVED: [Signature]
 Out of tolerance
 In tolerance
 Damaged (Contact customer)
 Contact _____ Date _____
 Received into system, introduced or reactivated

ACTION TAKEN: (Calibration/Repair/Parts) _____

Gage Rejected. Excessive Span error. (Probably due to over ranging) Sticking also around mid range
Replacement Suggested

CAL ENVIRONMENT: Temperature 68 °F Humidity 41 %RH

CALIBRATED/REPAIRED:
By [Signature] Cal Procedure CCCP-P1-001 Jun 99
Date 10-6-99 Accuracy Unknown
Cal Interval _____ Reliability Code _____
Next Cal Due _____ Cal Time 0.5h Repair Time _____
Standards used (Asset #) _____

DATE COMPLETED 10-6-99 Anne-Marie LEJEUNE

DATE PICKED UP 10-7-99 PICKED UP BY [Signature]

36231

