

6/2/90



CENTER FOR NUCLEAR WASTE REGULATORY ANALYSES QUALITY ASSURANCE SURVEILLANCE REPORT

PROJECT NO.: 20-3704-042

REPORT NO.: 90-003

PAGE 1 OF 2

SURVEILLANCE SCOPE:

Preparation of test solution and conduct of Cyclic Polarization Test # PTVR-14

REFERENCE DOCUMENTS: TOPs-010, 008

STARTING DATE: 5/28/90

ENDING DATE: 5/28/90

QA REPRESENTATIVE: R. D. Brient

PERSONS CONDUCTING TEST / EXAM / ACTIVITY:

W. Machowski (06)

SATISFACTORY FINDINGS:

See Attached.

UNSATISFACTORY FINDINGS:

Contrary to the requirement of TOP-008, para 6.1, a ASTM G-61 verification test was not performed before beginning the PTVR series of tests.

NONCONFORMANCE REPORT NO.:

90-003¹ NS

ATTACHMENTS:

Nonconformance Report No. 90-003¹ NS

RECOMMENDATIONS / ACTIONS

See Attached.

APPROVED: _____

Samuel Malin
CENTER DIRECTOR OF QUALITY ASSURANCE

DATE: _____

5/29/90

DISTRIBUTION:

ORIGINAL - CENTER QA DIRECTOR
ORIGINATOR
PRINCIPAL ENGINEER N. Sridhar
ELEMENT MANAGER P. Nair
A. Whiting

Surveillance Report 90-003

SATISFACTORY FINDINGS:

A. Preparation of Modified J-13 Water

For test PTVR-14, J-13 water modified with 20ppm. Cl⁻ was prepared as specified in TOP-010, paragraphs 7.3 and 8.1. Test solution pH and chloride concentration were measured using an Orion model EA 920 meter, calibrated before use with buffer (Fisher certified) and KCl solutions, respectively. The stock solutions A and B were prepared on an earlier date and labeled as specified in paragraphs 7.1 and 7.2. The stock solutions were within their expiration date of 6/6/90. Test solutions are being made daily, and are not being carried over. Required data was recorded in the Scientific Notebook.

B. Conduct of Test PTVR-14

1. The test specimen, Incoloy 825 Heat # HH 4371 FC, was machined to Drawing 20-3704-042 and TOP-003, and prepared by polishing and cleaning as specified in TOP-008, paragraph 6.2.1. Dimensions were measured (using micrometer s/n 20-M-1, calibrated 4/23/90 due 10/23/90) and surface area calculated.

2. The test apparatus utilized the EG&G PAR 273 Potentiostat and test cell as described in paragraph 5.1. The reference electrode was checked (using a Keithly 602 electrometer s/n 19831, calibrated 11/21/89 due 11/21/90) against a spare as required in paragraph 5.5 before and after the tests. Differences in potential were less than 3 millivolts.

3. The test cell was deaerated with high purity Argon for one hour before introducing the specimen. After another hour at 95 degrees C., the scanning started. Reverse scanning was initiated manually after approximately 2 1/2 hours. Data required by paragraph 6.2.3 was recorded in the Scientific Notebook.

RECOMMENDATIONS/ACTIONS

A. Test instrumentation used should be identified in the initial entries for the test series by brand, model, serial number, and calibration status. For instruments calibrated before use, such as the pH meter, the buffer and ion standard solutions should be identified by lot or solution i.d. number.

B. All lab prepared solutions, including ion standards and salt bridge solutions, should be labeled in the same fashion as described in TOP-010, paragraph 6. The test solution from test PTVR-13, which was held over from 5/25/90 for a final pH measurement, was stored in an unlabeled bottle, and is required by TOP-010 to be labeled properly. All prepared solutions should be documented by Scientific Notebook entries on their date of preparation, and include Lot numbers of their parent materials and a description of the method of preparation.

8/26



CENTER FOR NUCLEAR WASTE REGULATORY ANALYSES

DEVIATION AND NONCONFORMANCE REPORT

PAGE 1 OF 1

PURCHASE ORDER NO.		PROJECT NO. 20-3704-042	JOB REQUEST NO.	DNR NO. 90-008765		
ITEM NAME OR ACTIVITY Cyclic Polarization Tests; PTVR series			ACTIVITY PERFORMED BY IWPE/W. Machowski			
ITEM DESCRIPTION (S/N, MODEL, WELD, SITE, ELEMENT NAME, REGULATION, TEST, ETC.) Cyclic Polarization ASTM G-61 Verification Test						
DESCRIPTION OF DEVIATION OR NONCONFORMANCE Contrary to TOP-008, para 6.1, an ASTM G-61 verification test was not performed before the PTVR series of tests.						
PROBABLE CAUSE OF DEVIATION OR NONCONFORMANCE Unknown						
ORIGINATOR (NAME) R.D. Brient <i>[Signature]</i>			DATE 5/28/90			
ACTION TO CORRECT DEFICIENCY						
ACTION TO PREVENT RECURRENCE						
CORRECTIVE ACTION TO BE TAKEN BY (NAME)			TARGET DATE FOR INITIATION _____ OR COMPLETION _____			
DISPOSITION	ACCEPT <input type="checkbox"/>	REWORK <input type="checkbox"/>	SCRAP <input type="checkbox"/>	RETURN TO VENDOR <input type="checkbox"/>	HOLD <input type="checkbox"/>	OTHER (SPECIFY) <input type="checkbox"/>
BASIS FOR DECISION OR NOTES						
LIST OF ATTACHMENTS			STATUS: <input type="checkbox"/> OPEN <input type="checkbox"/> CLOSED			
ELEMENT MANAGER (SIGNATURE)		DATE	DISTRIBUTION: DIRECTOR OF QA ORIGINATOR ELEMENT MANAGER (PROJECT FILES) PERSON RESPONSIBLE FOR CORRECTIVE ACTION			
DIRECTOR OF QA (SIGNATURE)		DATE				