



CENTER FOR NUCLEAR WASTE REGULATORY ANALYSES  
QUALITY ASSURANCE  
SURVEILLANCE REPORT

PROJECT NO.: 20.3702.015	REPORT NO.: 92-10	PAGE 1 OF 1
SURVEILLANCE SCOPE: Review of CALIBRATION AND SHELF-LIFE STATUS		
REFERENCE DOCUMENTS: N/A		
STARTING DATE: 9/16/92	ENDING DATE: 9/16/92	
QA REPRESENTATIVE: R.D. BRIENT		
PERSONS CONDUCTING TEST / EXAM / ACTIVITY: SRIDHAR, PRIKEYL		
SATISFACTORY FINDINGS: Calibration due dates and shelf life expiration dates of reagents were checked in the Geochemistry and Corrosion Labs. Two instruments (not currently being used for measurements) were found with past due dates in the Corrosion Lab. One reagent was past its shelf life. All of these situations were immediately corrected.		
UNSATISFACTORY FINDINGS: None		
NONCONFORMANCE REPORT NO.: N/A		
ATTACHMENTS: None		
RECOMMENDATIONS / ACTIONS: Personnel should check shelf life dates each time reagents are used, and notify the P.I. when expired. Likewise, calibration due dates should be periodically checked to confirm the recall system.		
APPROVED: <u>R. D. Brient for P.I.</u> CENTER DIRECTOR OF QUALITY ASSURANCE	DISTRIBUTION: ORIGINAL - CENTER QA DIRECTOR ORIGINATOR PRINCIPAL ENGINEER SRIDHAR, PRIKEYL ELEMENT MANAGER NAIC, SABRE	
DATE: 9/30/92		