## CENTER FOR NUCLEAR WASTE REGULATORY ANALYSES DEVIATION AND NONCONFORMANCE REPORT

			PAGE 1 OF 2	
PURCHASE ORDER NO.	PROJECT NO.	JOB REQUEST NO		
	20-3704-022		90-003	
ITEM NAME OR ACTIVITY		ACTIVITY PERFOR	ACTIVITY PERFORMED BY	
Thermohydrology Research Project		S. Sveder	S. Svedeman, (04)	
ITEM DESCRIPTION (S/N, MODEL, WELD, SITE, ELEMENT NAME, REGULATION, TEST, ETC.)				
Separate Effects Experiment: Test #1 - Calibration of Densitometer				
DESCRIPTION OF DEVIATION OR NONCONFORMANCE				
See attached				
PROBABLE CAUSE OF DEVIATION OR NONCONFORMANCE				
PROBABLE GAUSE OF DEVIATION ON NONOCHI CHIMANOL				
Unfamiliarity with calibration systems requirements.				
ORIGINATOR (NAME)			DATE	
			2/0/00	
R.D. Brient Ly			3/9/90	
ACTION TO CORRECT DEFICIENCY				
ALUNINUM BLOCKS WILL BE MEASURED USING AN ACCEPTABLE				
MEASUREMENT TECHNIQUE				
ACTION TO PREVENT RECURRENCE  MEASUR ETENTS REQUIRING HIGHER LEVEL STANDARDS WILL BE				
NAPIE ACCORDING TO ACCEPTABLE METHODOLOGIES				
CORRECTIVE ACTION TO BE TAKEN BY (NAME)		TARGET DATE FO	TARGET DATE FOR INITIATION $3/26/90$	
5.5VEDENAN		OR COMPLETION	OR COMPLETION 4/9/90	
DISPOSITION ACCEP	T REWORK	SCRAP RETURN TO	VENDOR HOLD OTHER (SPEC	CIFY)
BASIS FOR DECISION OR NOT	TES .	•	.0.1.1.1.	
Use of the blocks as standards will be a coop table with the specified thickers verification				
LIST OF ATTACHMENTS			STATUS: Lik gladge	
		_	OPEN CLO	SED
ELEMENT MANAGER (SIGNAT	TURE) DATE	DISTRIBUTION:	•	
John Htu	use \$26/9	DIRECTOR OF CORIGINATOR	<i>I</i> A	
DIRECTOR OF QA (SIGNATURE) DATE ELEMENT N			GER (PROJECT FILES) DR. Russel	1
3/26/9 PERSON			RESPONSIBLE FOR CORRECTIVE ACTION	
Janu Mallilo " 1 RON GREEN				

8/32

## DNR 90-003 Description of Nonconformance

CQAM Section 12, "Control of Measuring and Test Equipment", requires that measurement standards have documented traceability to higher level standards. The notebook entries for the aluminum blocks used for the densitometer calibration do not provide an indication that their thickness ("calibration") was determined utilizing a calibrated dimensional measuring equipment.

## CENTER FOR NUCLEAR WASTE REGULATORY ANALYSES

MEMORANDUM

April 18, 1990

To:

DNR File 90-003

From:

Bob Brient

Subject: Close-out of DNR 90-003

Corrective action to close-out DNR 90-003 was verified based on the review of the Scientific Notebook entry addressing the nonconforming condition.

Aluminum blocks of various thicknesses were utilized to establish the response characteristics of the densitometer system through the range of interest. The initial thickness measurements of the blocks were made using an uncalibrated caliper, measuring to the nearest .001". Technical review determined that this level of accuracy was not necessary, and the blocks were remeasured with a scale (which does not require calibration) to the nearest .01"(a copy of the Scientific Notebook entry documenting this remeasurement is attached). The new measurements do not have any impact on the conclusions reached which were based on the initial measurements.

The exercise of correcting this nonconformance is adequate instruction to familiarize project personnel with instrument calibration requirements to preclude recurrence of the problem.

Approved by

Bruce Mabrito Director of QA

cc: R. Green

J. Russell

S. Svedeman (04)


Attachment

4-13-90

SINCE THE AL PLATE THICKNESS WERE MEASURED WITH AN "UN-TRACEABLE" CALIPER SET, THE MEASUREMENTER WERE REPEATED WITH A CRAFTSMAN 6" SCALE (#40896, WITH O.O.I" DIVISIONS. THE RESULS ARE:

BLOCK # 1 = 0.75", BLOCK # 2 = 0.75", BLOCK # 3 = 0.75"

BLOCK # 9 = 0.50", BLOCK # 6 = 0.76", BLOCK # 7 = 0.03"



THESE READING MATCH FAIRLY WELL WITH
THE PRIOR READINGS AND DO NOT
CHANGE THE RESCRETS IN ANY SIGNIFICANT
WAY. THE CALIBRATION CHECK VERIFIES
THE DENSITUMETER FUNCTIONS THROUGHOUT
THE RANGE OF INTEREST IN THE EXPECTES;
MANNER. Stee Sundam 4-13-90.