

U. S. ATOMIC ENERGY COMMISSION
BYPRODUCT MATERIAL LICENSE

Supplementary Sheet

37-2416-2

License Number (F61)

AMENDMENT NO. 15

Curtiss-Wright Corporation
Research Division
Nuclear Sciences and Engineering
Department
Quehanna, Pennsylvania

Attention: N. R. Whaclock, Chairman
Nuclear Hazards Committee

In accordance with letter dated February 23, 1960, and signed by M. T. Beam,
Condition 14 of License No. 37-2416-2 is amended to read as follows:

- 14. Sealed sources containing byproduct material (except solid metal Iridium 192, Tantalum 182, Cobalt 60 plated with gold or nickel, gases, or sources which are stored and used in the licensee's hot cells or storage pool only) shall be tested for leakage and contamination at intervals not exceeding six (6) months and records of test results shall be maintained by the licensee.

DUPLICATED
FOR DIV. OF INSP.

For the U. S. Atomic Energy Commission
Original Signed By
James R. Mason
Chief, Isotopes Branch

Date March 7, 1960

Division of Licensing and Regulation
Washington 25, D. C.

1. 26

REB 3/7/60

6-19