

**From:** Richard McKinley  
**To:** Rtokarz@optonline.net  
**Date:** 4/28/04 1:13PM  
**Subject:** St. Peter's renewal

Bob,

Two last questions:

- 1) For survey instrument calibration I need the statement "We have developed and will implement and maintain written survey meter calibration procedures in accordance with the requirements in 10 CFR 20.1501 and that meet the requirements of 10 CFR 35.61."
- 2) Please verify that the calibration of your Sr-90 eye applicator will meet 10 CFR 35.432 (a)(3) using currently accepted protocols.

Thank you for your efforts.

Sincerely,  
Richard McKinley  
Health Physicist  
NRC, RI

**CC:** Rwm1