CENTER FOR NUCLEAR WASTE REGULATORY ANALYSES

NONCONFORMANCE REPORT

Project No. 20.06002.01.081	NCR No. <u>2004-05</u>
PART 1: DESCRIPTION OF NONCONFORMANCE Thermometer Model #15-166A (Serial Number A2000-130, Ass normally scheduled calibration.	set number 008109) was found to be out of tolerance during a
Initiated by: Darrell S. Dunn	Date: 6/4/2004
PART 2: PROPOSED DISPOSITION AND CORRECTIVE	ACTION
Disposition: Accept data obtained in tests using this thermometer as is.	
Basis of Disposition: The thermometer had an error less than 1.0 degrees at 50, 100 at degrees. Thermometer was within calibration for the temperature thermometer were conducted in the range of 20 to 95 °C. Over 1 degree. An error in test temperatures of less than 2 °C will not	nd 150 °C. The thermometer should be accurate within 0.5 are range from -19 to 0 °C. Tests conducted with the this temperature range, the instrument error was less than t adversely affect the parameters measured during testing.
Action to Correct Nonconformance: Mercury separation that caused thermometer to be out of toleran tolerance.	ace was corrected and the thermometer was returned in
Proposed by: Darrell S. Dunn Dunel J.D.	Target date for completion: 6/11/2004 Date: 6/4/2004
PART 3: APPROVAL Element Manager: <u>Hypergener</u> Director of QA: <u>Upper</u> Comments/Instructions:	Date: 6/4/04 Date: 6/4/04
Comments: No further action required.	V. Jain B. Sagar
Verified by: Mach R. Ekustion Date: 6/11/04	

CH INSTITUT Drawer 28510 Systems Laboratory 210-522-3692								
	0972-01							
Work Ordori	44056800							
Nork Oruer: Data Issued:	Ian 8 2004							
Calibration Date:	Ian 8, 2004							
**Calibration Due:	Jul 8, 2004							
Calibration Location:	Bldg. 64							
Environment:	Temp. 76.0°F Hum. 33 %RH							
*As Found:	OUT OF TOLERANCE							
*As Left:	IN TOLERANCE							
conforms to ISO/IEC 17025, 1999 and ANSI/NCSL Z540-1-1994 which are equivalent to relevant requirements of the ISO 9000-1994 series of standards. This certificate may not be reproduced, except in full, without the written approval of the Southwest Research Institute Calibration Laboratory. The results of this calibration relate only to the individual instrument described above. This certificate shall not be used to claim product endorsement by the American Association for Laboratory Accreditation (A2LA) or any agency of the U. S. Government								
Guide to the Expression of Uncertainty ce level. The calibration process prov	in Measurement" (GUM). The uncertainty des a Test Uncertainty Ratio (TUR) of less							
*The client has sole responsibility for determination of in/out of tolerance or compliance/noncompliance. An in/out of tolerance opinion is provided for your convenience based only on the Test Instrument (TI) reading(s) and limits as reported. The reported uncertainty relates only to the results at the time of calibration and does not imply any short or long term stability of the TI. **Calibration interval is determined by the client and does not assure the instrument will remain within tolerance until this date. Any number of factors may cause the instrument to be out of tolerance before the next calibration date. Remarks: None								
Description	Cal Due							
THERMOMETER	Feb 05, 04							
	**Calibration Due: Calibration Location: 1 Environment: *As Found: 0 *As Found: 0 *As Left: 1 T) and the International System of Unit requirements of the ISO 9000-1994 set ibration Laboratory. The results of thi e American Association for Laboratory Guide to the Expression of Uncertainty ce level. The calibration process provi- bliance. An in/out of tolerance opinion to the results at the time of calibration to the results at the time of calibration within tolerance until this date. Any nu Description PLATINUM RTD THERMOMETER							

Watt chi

Approved by: Walt Hill Metrology Group Leader m:\a2la1.rpt Rev date 15, August 02

-

AA /a Kentz,

Measurements by: Mark Romero Metrology Technician

1 of 1

Southwest Research Institute Calibration laboratory Calibration Report

Work Order:	444056809	Mfr.	Fisher-Scientific	Technician	Mark Romerc
Asset No.	008109	Model	15-166A		
Serial No.	A2000-130	Туре.	THERMOMETER	Cal Date.	08-Jan-04
Remarks:	Corrected pre	evious mercur	y seperation, which caused or	ut of tolerance readin	gs.
					·····

_	Function/Range	Test Point	TI Reading	Difference	+/-Limit	+/-Uncertainty	Left	
	Temperature	Deg C	Deg C	Deg C	Deg C	Deg C	Result	
		-20.1	-19.9	0.2	0.5	0.6	Pass	
		0.1	0.3	0.2	0.5	0.6	Pass	
		50.0	49.9	-0.1	0.5	0.6	Pass	
		100.0	99.8	-0.2	0.5	0.6	Pass	
		150.0	149.8 END O	-0.2 F REPORT	0.5	0.6	Pass	

r

-



Southwest Research Institute Calibration laboratory Calibration Report

Work Order: Asset No.	444056809 008109	Mfr. Model	Fisher-Scientific 15-166A		Technician	Mark Romero
Serial No.	A2000-130	Туре.	THERMOMETER		Cal Date.	07-Jan-04
Remarks:					<u></u>	
Function/Range	Test Point	TI Reading	Difference	+/-Limit	+/-Uncertainty	Found
Temperature	Deg C	Deg C	Deg C	Deg C	Deg C	Result
	-19.9	-20.1	-0.2	0.5	0.6	Pass
	0.1	-0.2	-0.3	0.5	0.6	Pass
	50.0	49.3	-0.7	0.5	0.6	Fail
	100.0	99.1	-0.9	0.5	0.6	Fail
	149.9	149.0 END (-0.9 DF REPORT	0.5	0.6	Fail

-