



Southwest Research Institute
6220 Culebra Road
San Antonio, TX 78238
(210) 522-5215
Department of Quality Assurance
Calibration Laboratory

Accredited

Certificate #
0972-01

Certificate of Calibration

11 July 2000

Issued to: JIM PRIKRYL DIV20 B57
Manufacturer/Model: PARR 4843
Description: TEMPERATURE CONTROLLER
Serial Number: 2209
Asset Number: 007388

This certifies the above item was calibrated in compliance with MIL-STD-45662A and ANSI/NC SL Z540-1-1994. The results of this calibration relate only to the individual item as described above. Standards used in this calibration, described in the referenced calibration procedure with associated uncertainties or tolerances, are traceable to the National Institute of Standards and Technology (NIST). Supporting documentation relative to traceability is on file and available for examination upon request. This certificate is not to be reproduced, except in full, without the written approval of the Southwest Research Institute Department of Quality Assurance Calibration Laboratory.

This laboratory is accredited by the American Association for Laboratory Accreditation (A2LA) and the results of this calibration certificate were determined in accordance with the terms of accreditation unless stated otherwise below.

The uncertainty of the calibration was sufficient to determine that the item met the manufacturer's published specifications unless stated otherwise below.

Ambient Conditions: Temperature: 68.0 Degrees Fahrenheit Humidity: 46 % RH

Calibration Date: 11 Jul 00 **Calibration Procedure:** CL100 6/99

Condition as Received: In Tolerance


Remarks:

Approved by:



Jim Patterson, Supervisor, or Walt Hill, Metrologist

Measurements performed by:



Mack Wood, Technician

Certificate # 39665
m\la2la.rpt Rev date 22 May 00

Page 1 of 1