

U.S. NUCLEAR REGULATORY COMMISSION		Conversation Date: 10/21/2004	
TELEPHONE CONVERSATION RECORD		Time: 1540	
Mail Control No.:	License No.:	Docket No.:	
135724	37-05809-02	03006135	
Licensee/Applicant Participant(s):	Organization:	Telephone No.:	
Ms. Lisa Lamantia, RSO	Curtiss-Wright Electro-Mechanical Corporation	724 275-5032	
Person(s) Calling: Ron Rolph			
Subject: Request for additional information			
<p>Summary:</p> <p>The application requests that a x-ray Fluorescence device be add to the current license. In the supplemental provided, NUREG 1556 Volume 1 is referenced. This is the incorrect volume. The references should be to volume 7.</p> <p>The RSO will resend the supplemental information as required by NUREG 1556 Volume 7 Appendix D.</p> <p>I also spoked to the RSO concerning use of the device at other locations and the need to change NRC Form 313 item 3 if temporary locations are to be used.</p> <p>The RSO will check with the engineering manager and verify that the device is only to be used at their facility. If it is to be used else where, the RSO will send a change to item 3 of NRC Form 313.</p>			
Action Required/Taken: Document			
Prepared By: Ron Rolph		Date: 10/21/04	