

U.S. NUCLEAR REGULATORY COMMISSION		Conversation Date: 11/15-19/04	
TELEPHONE CONVERSATION RECORD		Time: various	
Mail Control No.:	License No.:	Docket No.:	
135808	37-13831-01	03003203	
Licensee/Applicant Participant(s):	Organization:	Telephone No.:	
Michael Erdman	Milton S. Hershey Medical Center	717-531-8574	
Person(s) Calling: SGabriel, RI			
Subject: License amendment request to add Am-241 calibration sources			
<p>Summary:</p> <p>This record is to represent discussions with Mike Erdman during the on-site inspection on 11/15-17/04 and e-mail communication on 11/19/04.</p> <p>Licensee requested to add NIST-traceable disc sources of Am-241 for use in instrument calibration, with a maximum possession limit of 15 microcuries. I informed Mike that it is not necessary to specifically list the sources on the license if they are covered by 10 CFR 35.65. 35.65(d) allows possession of any byproduct material with a half-life longer than 120 days in individual amounts not to exceed the smaller of 200 uCi or 1000 times the quantities in Appendix B of 10 CFR Part 30. We reviewed Appendix B and determined that 0.01 uCi is the value for Am-241. 1000 times this value is 10 uCi. Mike stated that his intention was for each source to be no greater than 10 uCi. Mike asked to withdraw the request to specifically list Am-241 calibration sources on the license.</p>			
Action Required/Taken: Issue amendment without Am-241 sources; explain in cover letter			
Prepared By: SGabriel		Date: 11/27/04	