

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED: Milton S. Hershey Medical Center Hershey, Pennsylvania 17033		2. NRC/REGIONAL OFFICE U.S. Nuclear Regulatory Commission Region I, 475 Allendale Road King of Prussia, Pennsylvania 19406-1415	
REPORT NUMBER(S) 2004001			
3. DOCKET NUMBER(S) 03003203/03010851	4. LICENSEE NUMBER(S) 37-13831-01/37-13831-04	5. DATE(S) OF INSPECTION November 16-18, 2004	

LICENSEE:

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:

- 1. Based on the inspection findings, no violations were identified.
- 2. Previous violation(s) closed.
- 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, NUREG-1600, to exercise discretion, were satisfied.

_____ Non-Cited Violation(s) was/were discussed involving the following requirement(s) and Corrective Action(s):

- 4. During this inspection certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11.

(Violations and Corrective Actions)

10 CFR 20.1801 requires that the licensee secure from unauthorized removal or access licensed materials that are stored in controlled or unrestricted areas. 10 CFR 20.1802 requires that the licensee control and maintain constant surveillance of licensed material that is in a controlled or unrestricted area and that is not in storage. As defined in 10 CFR 20.1003, *controlled area* means an area, outside of a restricted area but inside the site boundary, access to which can be limited by the licensee for any reason; and *unrestricted area* means an area, access to which is neither limited nor controlled by the licensee.

Contrary to the above, on November 21, 2003, the licensee did not secure from unauthorized removal or limit access to a trash bag containing 26 microcuries of phosphorus-32 located in a research laboratory, which is normally a controlled area, nor did the licensee control and maintain constant surveillance of this licensed material.

Corrective actions included provision of additional training for Environmental Health staff members and initiation of use of yellow trash bags for all radioactive waste.

Licensee's Statement of Corrective Actions for Item 4, above.

I hereby state that, within 30 days, the actions described by me to the inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

Title	Printed Name	Signature	Date
LICENSEE'S REPRESENTATIVE			
NRC INSPECTOR	Sandra Gabriel/Tara Weidner	RA/RA	11/29/04