

**SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION**

1. LICENSEE/LOCATION INSPECTED: Internal Medicine Associates of Mt. Clemens, P.C. 133 South Main Street, Suite A Mt. Clemens, Michigan 48043  REPORT                  2004-002, -003	2. NRC/REGIONAL OFFICE  REGION III US NUCLEAR REGULATORY COMMISSION 2443 WARRENVILLE ROAD, SUITE 210 LISLE, ILLINOIS 60532
---	---

3. DOCKET NUMBER(S) 030-33050	4. LICENSEE NUMBER(S) 21-26470-01	5. DATE(S) OF INSPECTION 11/19/04
----------------------------------	--------------------------------------	--------------------------------------

**LICENSEE:**

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:

- 1. Based on the inspection findings, no violations were identified.
- 2. Previous violation(s) closed.
- 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, NUREG-1600, to exercise discretion, were satisfied.

\_\_\_\_\_ Non-Cited Violation(s) was/were discussed involving the following requirement(s) and Corrective Action(s):

- 4. During this inspection certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11.

(Violations and Corrective Actions)

**Licensee's Statement of Corrective Actions for Item 4, above.**

I hereby state that, within 30 days, the actions described by me to the inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

Title	Printed Name	Signature	Date
LICENSEE'S REPRESENTATIVE			
NRC INSPECTOR	Geoffrey M. Warren		11/19/04

**Docket File Information**  
**SAFETY INSPECTION REPORT**  
**AND COMPLIANCE INSPECTION**

1. LICENSEE <b>Internal Medicine Assoc. Of Mt. Clemens, P.C.</b> REPORT NUMBER(S) 2004-002, -003		2. NRC/REGIONAL OFFICE <b>Region III</b>	
3. DOCKET NUMBER(S) <b>030-33051</b>	4. LICENSE NUMBER(S) <b>21-26470-01</b>	5. DATE(S) OF INSPECTION <b>11/19/04</b>	
6. INSPECTION PROCEDURES USED <b>87130</b>	7. INSPECTION FOCUS AREAS <b>03.01 - 03.07</b>		

**SUPPLEMENTAL INSPECTION INFORMATION**

1. PROGRAM CODE(S) <b>02201</b>	2. PRIORITY <b>5</b>	3. LICENSEE CONTACT <b>Gunateet K. Goswami, M.D., RSO</b>	4. TELEPHONE NUMBER <b>586-465-1326</b>
------------------------------------	-------------------------	--	--

Main Office Inspection      Next Inspection Date: **Nov. 2009**

Field Office \_\_\_\_\_

Temporary Job Site \_\_\_\_\_

**PROGRAM SCOPE**

The licensee was a nuclear medicine clinic located in Mt. Clemens, Michigan. Licensee had authorization to use byproduct materials under 10 CFR 35.100 and 35.200. Licensed activities were conducted at the location indicated on the license. The nuclear medicine department was staffed with one full-time nuclear medicine technologist and one fill-in technologist who came in as needed. The licensee's nuclear medicine staff typically administered 200 diagnostic doses monthly. Doses were primarily technetium-99m for cardiac studies with occasional bone, hepatobiliary, and other studies. Doses were received as unit doses from a licensed radiopharmacy. All waste was either held for decay-in-storage (DIS) or returned to the radiopharmacy.

**Performance Observations**

The inspector observed one diagnostic administration of licensed material, including dose preparation and disposal, and identified no concerns with the procedures. Licensee staff demonstrated package receipt and return surveys, daily dose calibrator and survey meter constancy checks, and daily and weekly contamination surveys. The inspector found no concerns with these activities. Interviews with licensee staff indicated adequate knowledge of radiation safety procedures. Surveys indicated radiation levels consistent with licensee radiation survey records and postings.

