

# Region I NMSS Licensee Event Report

Licensee	AGILENT TECHNOLOGIES
Event Description	NOTICE OF GENERAL LICENSE DEVICE w/HIGH REMOVABLE ACTIVITY

License #	07-28762-02G	Docket #	03002988	MLER-RI	2004-05/
Event Date	MAY 04	Report Date			

1. REPORTING REQUIREMENT

<input type="checkbox"/>	10 CFR 20.2201 Theft or Loss	<input type="checkbox"/>	10 CFR 35.33 Misadministration
<input type="checkbox"/>	10 CFR 20.2203 30 Day Report	<input type="checkbox"/>	License Condition
<input type="checkbox"/>	10 CFR 30.50 Report		
XXX	Other 10 CFR 31.5		

2. REGION I RESPONSE

<input type="checkbox"/>	Immediate Site Inspection	Inspector/Date	<input type="text"/>
<input type="checkbox"/>	Special Inspection	Inspector/Date	<input type="text"/>
<input type="checkbox"/>	Telephone Inquiry	Inspector/Date	<input type="text"/>
<input type="checkbox"/>	Preliminary Report	<input type="checkbox"/>	Daily Report
XX	Information Entered in RI Log	<input checked="" type="checkbox"/>	Review at next inspection
<input type="checkbox"/>	Report referred to:	<input type="text"/>	

3. REPORT EVALUATION

<input checked="" type="checkbox"/>	Description of Event	<input checked="" type="checkbox"/>	Corrective Actions
<input checked="" type="checkbox"/>	Levels of RAM Involved	<input type="checkbox"/>	Calculations Adequate
<input checked="" type="checkbox"/>	Cause of Event	<input type="checkbox"/>	Additional Information Requested from Licensee

4. SPECIAL INSTRUCTIONS OR COMMENTS

<i>Multiple Documents</i>	5/4/04
	5/10/04
	5/13/04
	5/17/04


Public ☒

Non-Public ☐

Completed by:	<i>E. Ullrich</i>	Date	<i>BY 9/13/04</i>
Reviewed by:	<i>Joe D. Kenna</i>	Date	<i>9.16.2004</i>

Location of File: G:\LAS\LER FORM REV A.wpd

ML043430028

 **Agilent Technologies**  
2850 Centerville Road  
Wilmington, DE 19808

RECEIVED  
REGION 1

'04 MAY -6 P1:31

Division of Nuclear Materials Safety  
U.S. Nuclear Regulatory Commission  
Region I  
475 Allendale Road,  
King of Prussia, PA 19406

Re: Notice of General License device with high removable activity (NRC License 07-28762-02G)

May 4, 2004

Dear Sir or Madam,

Per the requirements of 10 CFR 31.5, I am notifying your office of a general license device from which we have obtained a removable activity wipe in excess of 0.005  $\mu\text{Ci}$ . The cell was evaluated by support engineering and its disposition is identified below.


Serial #	Model #	Activity ( $\mu\text{Ci}$ )	Disposition
F1909	19233	0.034	Scrapped


Comments:

The cell was used for internal instruments testing for 10 years. The source was corroded and very dark.

Please contact me at 302-633-8262 if there are any questions.

Thank you,

  
David S. Bennett  
Radiation Safety Officer

 **Agilent Technologies**  
2850 Centerville Road  
Wilmington, DE 19808

RECEIVED  
REGION 1

'04 MAY 11 P1:01

Division of Nuclear Materials Safety  
U.S. Nuclear Regulatory Commission  
Region I  
475 Allendale Road,  
King of Prussia, PA 19406

Re: Notice of General License device with high removable activity (NRC License 07-28762-02G)

May 10, 2004

Dear Sir or Madam,

Per the requirements of 10 CFR 31.5, I am notifying your office of a general license device returned to our site by a customer from which we have obtained a removable activity wipe in excess of 0.005  $\mu\text{Ci}$ . The cell was evaluated by support engineering and its disposition is identified below.


Serial #	Model #	Activity ( $\mu\text{Ci}$ )	Disposition
F4461	19233	0.033	Scrapped


Comments:

The cell was 12 years old. The source was dirty, dark, and corroded.

Please contact me at 302-633-8262 if there are any questions.

Thank you,

  
David S. Bennett  
Radiation Safety Officer

 **Agilent Technologies**  
2850 Centerville Road  
Wilmington, DE 19808

RECEIVED  
REGION 1

'04 MAY 17 P2:06

Division of Nuclear Materials Safety  
U.S. Nuclear Regulatory Commission  
Region I  
475 Allendale Road,  
King of Prussia, PA 19406

Re: Notice of General License device with high removable activity (NRC License 07-28762-02G)

May 13, 2004


Dear Sir or Madam,

Per the requirements of 10 CFR 31.5, I am notifying your office of the following devices returned to us by customers from which we have obtained removable activity wipes in excess of 0.005  $\mu\text{Ci}$ . The cells were evaluated by support engineering and their disposition is identified below.


Serial #	Model #	Activity ( $\mu\text{Ci}$ )	Disposition	Comments
F1083	19233	0.013	Scrapped	Cell was 13 years old.
F6118	19233	0.026	Scra[ed	Cell was 11 years old. Source was dirty and very dark.

Please contact me at 302-633-8262 if there are any questions.

Thank you,



David S. Bennett  
Radiation Safety Officer

 **Agilent Technologies**  
2850 Centerville Road  
Wilmington, DE 19808

RECEIVED  
REGION 1

'04 MAY 18 A11 :07

Division of Nuclear Materials Safety  
U.S. Nuclear Regulatory Commission  
Region I  
475 Allendale Road,  
King of Prussia, PA 19406

Re: Notice of General License device with high removable activity (NRC License 07-28762-02G)

May 17, 2004

Dear Sir or Madam,

Per the requirements of 10 CFR 31.5, I am notifying your office of a general license device returned to our site by a customer from which we have obtained a removable activity wipe in excess of 0.005  $\mu\text{Ci}$ . The cell was evaluated by support engineering and its disposition is identified below.

Serial #	Model #	Activity ( $\mu\text{Ci}$ )	Disposition
F3222	19233	0.006	Scrapped

Comments:

The cell was 13 years old. The source was rusted and dirty.

Please contact me at 302-633-8262 if there are any questions.

Thank you,

  
David S. Bennett  
Radiation Safety Officer