

# Region I NMSS Licensee Event Report

Licensee	AGILENT TECHNOLOGIES
Event Description	NOTICE OF GENERAL LICENSE DEVICE w/HIGH REMOVABLE ACTIVITY

License #	07-28762-02G	Docket #	03002988	MLER-RI	2004- 049
Event Date	March 04	Report Date			

1. REPORTING REQUIREMENT

<input type="checkbox"/> 10 CFR 20.2201 Theft or Loss	<input type="checkbox"/> 10 CFR 35.33 Misadministration
<input type="checkbox"/> 10 CFR 20.2203 30 Day Report	<input type="checkbox"/> License Condition
<input type="checkbox"/> 10 CFR 30.50 Report	
<input checked="" type="checkbox"/> Other 10 CFR 31.5	

2. REGION I RESPONSE

<input type="checkbox"/> Immediate Site Inspection	Inspector/Date	
<input type="checkbox"/> Special Inspection	Inspector/Date	
<input type="checkbox"/> Telephone Inquiry	Inspector/Date	
<input type="checkbox"/> Preliminary Report	<input type="checkbox"/> Daily Report	
<input checked="" type="checkbox"/> Information Entered in RI Log	<input checked="" type="checkbox"/> Review at next inspection	
<input type="checkbox"/> Report referred to: _____		

3. REPORT EVALUATION

<input checked="" type="checkbox"/> Description of Event	<input checked="" type="checkbox"/> Corrective Actions
<input checked="" type="checkbox"/> Levels of RAM Involved	<input type="checkbox"/> Calculations Adequate
<input checked="" type="checkbox"/> Cause of Event	<input type="checkbox"/> Additional Information Requested from Licensee

4. SPECIAL INSTRUCTIONS OR COMMENTS

<i>Multiple Documents</i>	<i>3/4/04</i>	<i>3/31/04</i>
	<i>3/18/04</i>	
	<i>3/25/04</i>	

Public  Non-Public

Completed by:	<i>E. ULLRICH</i> <i>EU</i>	Date	<i>9/13/04</i>
Reviewed by:	<i>[Signature]</i>	Date	<i>9.16.2004</i>

 **Agilent Technologies**  
2850 Centerville Road  
Wilmington, DE 19808

RECEIVED  
REGION 1

'04 MAR -5 P12 :36

Division of Nuclear Materials Safety  
U.S. Nuclear Regulatory Commission  
Region I  
475 Allendale Road,  
King of Prussia, PA 19406

Re: Notice of General License device with high removable activity (NRC License 07-28762-02G)

March 4, 2004

Dear Sir or Madam,

Per the requirements of 10 CFR 31.5, I am notifying your office of a general license device returned to our site by a customer from which we have obtained a removable activity wipe in excess of 0.005  $\mu\text{Ci}$ . The cell was evaluated by support engineering and its disposition is identified below.


<u>Serial #</u>	<u>Model #</u>	<u>Activity (<math>\mu\text{Ci}</math>)</u>	<u>Disposition</u>
U2667	G2397A	0.015	scrapped

Comments:

The cell was 4 years old.

Please contact me at 302-633-8262 if there are any questions.

Thank you,

  
**David S. Bennett**  
Radiation Safety Officer

 **Agilent Technologies**  
2850 Centerville Road  
Wilmington, DE 19808

RECEIVED  
REGION 1

'04 MAR 19 P1:13

Division of Nuclear Materials Safety  
U.S. Nuclear Regulatory Commission  
Region I  
475 Allendale Road,  
King of Prussia, PA 19406

Re: Notice of General License device with high removable activity (NRC License 07-28762-02G)

March 18, 2004

Dear Sir or Madam,

Per the requirements of 10 CFR 31.5, I am notifying your office of the following devices returned to us by customers from which we have obtained removable activity wipes in excess of 0.005  $\mu\text{Ci}$ . The cells were evaluated by support engineering and their disposition is identified below. All cells were returned from overseas customers.

Serial #	Model #	Activity ( $\mu\text{Ci}$ )	Disposition	Comments
L4098	19233	0.264	Scrapped	Source dark and dirty, cell was 15 years old
L6683	19233	0.039	Scrapped	Source rusted and dirty, cell was 15 years old
L4112	19233	0.023	Scrapped	Source dark and corroded, cell was 15 years old

Please contact me at 302-633-8262 if there are any questions.

Thank you,

  
David S. Bennett  
Radiation Safety Officer

 **Agilent Technologies**  
2850 Centerville Road  
Wilmington, DE 19808

RECEIVED  
REGION 1

'04 MAR 30 A7:08

Division of Nuclear Materials Safety  
U.S. Nuclear Regulatory Commission  
Region I  
475 Allendale Road,  
King of Prussia, PA 19406

Re: Notice of General License device with high removable activity (NRC License 07-28762-02G)

March 25, 2004

Dear Sir or Madam,

Per the requirements of 10 CFR 31.5, I am notifying your office of a general license device returned to our site by a customer from which we have obtained a removable activity wipe in excess of 0.005  $\mu\text{Ci}$ . The cell was evaluated by support engineering and its disposition is identified below.


<u>Serial #</u>	<u>Model #</u>	<u>Activity (<math>\mu\text{Ci}</math>)</u>	<u>Disposition</u>
L2623	19235	0.008	Scrapped

Comments:

The cell was 17 years old. The source was dirty and dark.

Please contact me at 302-633-8262 if there are any questions.

Thank you,

  
David S. Bennett

Radiation Safety Officer

 **Agilent Technologies**  
2850 Centerville Road  
Wilmington, DE 19808

RECEIVED  
REGION 1

'04 APR -1 P 1 :22

Division of Nuclear Materials Safety  
U.S. Nuclear Regulatory Commission  
Region I  
475 Allendale Road,  
King of Prussia, PA 19406

Re: Notice of General License device with high removable activity (NRC License 07-28762-02G)

March 31, 2004

Dear Sir or Madam,

Per the requirements of 10 CFR 31.5, I am notifying your office of the following devices returned to us by customers from which we have obtained removable activity wipes in excess of 0.005  $\mu\text{Ci}$ . Cell K1354 was returned from a U.S. customer and YA349 from a foreign one. The cells were evaluated by support engineering and their disposition is identified below.

Serial #	Model #	Activity ( $\mu\text{Ci}$ )	Disposition	Comments
K1354	G1533A	0.015	Transferred to new location after cleaning.	Cell was 9 years old
YA349	G1533A	0.012	Scrapped	Source burnt

Please contact me at 502-633-8262 if there are any questions.

Thank you,

  
David S. Bennett

Radiation Safety Officer

*[Faint, mostly illegible text at the bottom of the page, likely a copy of the letter's body.]*