

From: Penny Lanzisera
To: rojgm1@carilion.com
Date: Fri, Nov 26, 2004 3:01 PM
Subject: Amendment Request - Request for Additional Information
Place: LAT@nrc.gov

Licensee: Carilion Health System
RSO: Jeffrey G. Messinger
Mail Control No: 135674
License No. : 45-25395-01
Docket No.: 030-34470

Subject: Request for Additional Information

Dear Mr. Messinger, please provide:

- 1) Disposal documentation for your Cordis and Guidant intravascular brachytherapy sources, including confirmation of receipt by the manufacturer or waste broker.
- 2) The names of the authorized users (e.g. Drs. Heath, Scruggs, & Hess) and authorized medical physicists who will perform source calibrations and spot-checks (e.g., Mssrs. Surace & Kovats) that you are requesting authorization for the Novoste intravascular brachytherapy device and confirm that these individuals will have vendor training **prior** to use of the device.
- 3) Confirm that your dosimetry system meets the requirements of 10 CFR 35.630(a).

Please reference Mail Control No. 135674 in your written response. You may fax the response to 610-337-5269. If you have any questions, please call me at 610-337-5169.

Sincerely, Penny Lanzisera
Senior Health Physicist, NRC Region 1

CC: LAT@nrc.gov

Mail Envelope Properties (41A78B92.270 : 1 : 35540)

Subject: Amendment Request - Request for Additional Information
Creation Date: Fri, Nov 26, 2004 3:01 PM
From: Penny Lanzisera

Created By: PAN@nrc.gov

Recipients

carilion.com
rojgm1 (rojgm1@carilion.com)

nrc.gov
kp1_po.KP_DO
LAT CC (LAT@nrc.gov)

Post Office

kp1_po.KP_DO

Route

carilion.com
nrc.gov

Files

MESSAGE

Size

2316

Date & Time

Friday, November 26, 2004 3:01 PM

Options

Expiration Date: None
Priority: Standard
Reply Requested: No
Return Notification: None

Concealed Subject: No
Security: Standard