

U.S. NUCLEAR REGULATORY COMMISSION		Conversation Date	12/10/04
<b>TELEPHONE CONVERSATION RECORD</b>		Time	2:30 p.m.
Mail Control No.:	License No.:	Docket No.:	
135973	47-06773-02	030-14874	
Licensee/Applicant Participant(s):	Organization:	Telephone No.:	
Frank Hinricks, RSO	Bayer Material Sc.	304 451 2592	
Person(s) Calling: Sattar Lodhi			
Subject: Renewal Application			
<p>Summary:</p> <p>The licensee's renewal application did include procedures for routine and non routine maintenance of the gauges. I called the RSO to request the information and suggested that he review appendix B of NUREG 1556, Volume 4 and use the language that is suggested in the routine and non-routine maintenance section.</p> <p>He will have the requested information sent as soon as possible. Gave him RI fax number and control number.</p>			
Action Required/Taken: Document			
Prepared By: Sattar Lodhi		Date: 12/10/04	