



The  
**Cardiac Imaging**  
Center

1703 South Broad Street  
Suite 400  
Philadelphia, PA 19148  
(215) 339-0409

March 8, 1997

Nuclear Medicine Safety Branch  
Division of Radiation Safety and Safeguards  
U.S. Nuclear Regulatory Commission, Region I  
475 Allendale Road  
King of Prussia, PA 19406-1415

RE: License Number: 37-28653-01  
Amendment Application

Dear License Reviewer:

Please amend the above referenced byproduct material license to add a "location of use". The new facility is located at:

Cardiac Imaging Center  
Makefield Executive Offices  
301 Oxford Valley Road, Unit 901  
Yardley, PA 19067

Veronica Covalesky, M.D. will remain as Radiation Safety Officer. A diagram of the new department has been enclosed within "Attachment A" for your examination. All remaining conditions established within the license application and byproduct material license will be employed at this institution. Within "Attachment B" an "Equipment List" has been enclosed for your review.

This site is currently operated by NMC Diagnostic Imaging Services. An amendment application will be submitted by representatives at NMC Diagnostic Imaging Services. This request the deletion of this facility as a "location of use" on their byproduct material license.

Sealed sources used by NMC will be transferred to a licensed recipient. All sources will be transported in compliance with DOT regulations. Documentation regarding this transfer will be placed on file for future review.

**NMSS/RGNI MATERIALS-002**

Through conversations with representatives at the Nuclear Regulatory Commission, it has been determined that a "Close-Out Survey" will not be required. We request the imaging center be deleted from the NMC license and added as a "location of use" on the byproduct materials license issued to Cardiac Imaging Center. We request this licensing modification occur with no interruption in nuclear medicine operations.

We do not intend to utilize Iodine-131 for diagnostic and/or therapy applications. Therefore, a "Quality Management Program" will not be employed.


In regards to our imaging center located at 700 Cottman Avenue, Philadelphia, PA, we request authorization be granted to include a room where patients may be administered the radiopharmaceutical for diagnostic testing. A diagram has been enclosed within "Attachment C" indicating the "injection room. Please refer to this diagram for specific details.

The patient injection room will be added to the daily and weekly survey logs. The room will be evaluated on a daily basis for ambient exposure contamination and weekly for the assessment of removable contamination. A "radioactive materials" sign will be displayed on the door.

If you have additional questions, please contact Michael W. Lairmore or myself. Mr. Lairmore may be contacted at (201) 447-3303.

We thank you in advance for your assistance with this licensing action.

Sincerely,

A handwritten signature in black ink, appearing to read "Mark Victor, M.D.", with a large, sweeping flourish at the end.

Mark Victor, M.D.  
President

enclosures