

From: Michelle Beardsley
To: Dallan@radiology-inc.com
Date: Fri, Dec 10, 2004 11:31 AM
Subject: License renewal letter
Place: LAT@nrc.gov

M. Douglass Allan, M.S.
Putnam General Hospital
License No. 47-23070-01
Docket No. 03020237

Dear Mr. Allan:

This is in reference to your application dated July 29, 2004 requesting to renew Nuclear Regulatory Commission License No. 47-23070-01. In order to continue our review, we need the following additional information:

1. In your application you committed to establishing and implementing many of the Model Procedures published in NRC Regulatory Guide 10.8, Revision 2, "Guide for the Preparation of Applications for Medical Use Programs." This guide was published in August 1987 and was designed for use with the NRC's medical regulations that existed in 10 CFR Part 35 at that time. On April 24, 2002, NRC published new medical regulations in 10 CFR Part 35 (Enclosure 1). These regulations became effective on October 24, 2002. Concurrent with the issuance of the new medical regulations, NRC published NUREG-1556, Volume 9, "Consolidated Guidance About Materials Licenses: Program-Specific Guidance About Medical Use Licenses" (Enclosure 2). As specified in Appendix AA of NUREG-1556, Volume 9, Regulatory Guide 10.8, Revision 2 is superceded and should no longer be used.

Please use NUREG-1556, Volume 9, in preparing your license renewal application. You will note that new Part 35 and this guidance generally do not require the submission of detailed procedures during the licensing process. As detailed in NUREG-1556, Vol. 9, in many cases a licensee is required only to supply a statement regarding the development, implementation, and maintenance of written operating and emergency procedures. Appendix C of NUREG-1556, Volume 9, should be helpful in identifying the information required by NRC to process your request for license renewal. You may access this NUREG on our website at www.nrc.gov under Nuclear Materials >Quicklinks>Guidance>NUREG series.

In addition, please note that the authorized users that you have listed to use 35.300 materials for **diagnostic use only** will be listed on your renewed license to use "oral administration of sodium iodide I-131 for imaging and localization studies".

If you have any additional questions, you may call me at 610-337-6942. You may respond to this email by fax at 610-337-5393 or mail (not both). Please be sure to put the Control No. listed above on your response.

Sincerely,
Michelle Beardsley
Health Physicist
USNRC RI
475 Allendale Road
King of Prussia, PA 19406

CC: LAT@nrc.gov

Mail Envelope Properties (41B9CF76.36A : 12 : 35316)

Subject: License renewal letter
Creation Date: Fri, Dec 10, 2004 11:31 AM
From: Michelle Beardsley

Created By: MRB@nrc.gov

Recipients

nrc.gov
kp1_po.KP_DO
LAT CC (LAT@nrc.gov)

radiology-inc.com
Dallan (Dallan@radiology-inc.com)

Post Office
kp1_po.KP_DO

Route
nrc.gov
radiology-inc.com

Files	Size	Date & Time
MESSAGE	3778	Friday, December 10, 2004 11:31 AM

Options

Expiration Date: None
Priority: Standard
Reply Requested: No
Return Notification: None

Concealed Subject: No
Security: Standard