

Docket File Information
SAFETY INSPECTION REPORT
AND COMPLIANCE INSPECTION

1. LICENSEE Advanced Service Providers REPORT NUMBER(S) 2004-002/2004-003		2. NRC/REGIONAL OFFICE Region III 2443 Warrenville Road Lisle, IL 60532	
3. DOCKET NUMBER(S) 030-36444	4. LICENSE NUMBER(S) 48-32464-01	5. DATE(S) OF INSPECTION December 1, 2004	
6. INSPECTION PROCEDURES USED 87124	7. INSPECTION FOCUS AREAS 03.01-03.07		

SUPPLEMENTAL INSPECTION INFORMATION

1. PROGRAM CODE(S) 3225	2. PRIORITY K5	3. LICENSEE CONTACT Steve Conour, RSO	4. TELEPHONE NUMBER 515-314-9962 (Cell)
<input type="checkbox"/> Main Office Inspection <input checked="" type="checkbox"/> Field <u>Inspection Conducted at the Iowa Department of Public Health, Des Moines, IA</u> <input type="checkbox"/> Temporary Job Site			

Next Inspection Date: **December 2005**

PROGRAM SCOPE

This was an initial inspection of this fixed gauge service provider with a corporate location in Green Bay, WI.

The licensee performs service, maintenance, leak testing and installation of fixed gauges for approximately 26 clients in 12 states. About 20 authorized technicians are employed and service the devices at the various locations. Currently, the licensee does not possess licensed material as authorized on the current NRC license.

According to the RSO, 90 percent of the current client base possess generally licensed devices containing krypton-85.

Since gauge work was not being performed in the Des Moines, IA area, the inspection included interviews with the RSO as well as reviews of related documentation.

A random sample of technician training/experience records were reviewed and appeared adequate. In addition, emergency and security procedures were reviewed with the licensee as well as the licensee's radiation safety manual with no problems noted.

One gauge installation has been performed to date by the licensee. Although records were unavailable, the RSO indicated that surveys following gauge installation had been performed by the service technician and by the client.

A random sample of client service records were reviewed and did not indicate improper or unauthorized activities. Calibrated survey instruments are available through the licensee or at the client location.

Personal dosimetry is provided to each technician and exchanged quarterly. A review of dosimetry reports for 2003 and YTD 2004 did not indicate exposures in excess of 10 CFR 20 limits. According to the RSO, service was provided to gauge receivers/detectors only during 2003, and did not include work involving the actual gauge containing the source.

NOTE: Since byproduct material is not currently possessed, another inspection should be conducted in one year as stated above.

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED:

Advanced Service Providers
3014 White Sands Terrace
Green Bay, WI 54313

2. NRC/REGIONAL OFFICE

U.S. Nuclear Regulatory Commission
Region III
2443 Warrenville Road
Lisle, Illinois 60532-4351

REPORT

2004-001

3. DOCKET NUMBER(S)

030-36444

4. LICENSEE NUMBER(S)

48-32464-01

5. DATE(S) OF INSPECTION

~~November~~ 1, 2004
December

LICENSEE:

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:



1. Based on the inspection findings, no violations were identified.



2. Previous violation(s) closed.



3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, NUREG-1600, to exercise discretion, were satisfied.

Non-Cited Violation(s) was/were discussed involving the following requirement(s) and Corrective Action(s):



4. During this inspection certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11.

(Violations and Corrective Actions)

Licensee's Statement of Corrective Actions for Item 4, above.

I hereby state that, within 30 days, the actions described by me to the inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

Title

Printed Name


Signature

Date

LICENSEE'S
REPRESENTATIVE

NRC INSPECTOR

S. J. Mulay



12/1/04