

BRANSCOME INC.

Post Office Drawer 260
Williamsburg, Virginia 23187

WILLIAMSBURG: (757) 229-2504
NORFOLK: (757) 622-4200
FAX: (757) 220-0390

December 22, 2004

Control # 135777

Mr. Ron Rolph
Nuclear Regulatory Commission
Atlanta Federal Center
61 Forsyth Street, SW, Suite 23T85
Atlanta, GA 30323

MS 16

Q-7

45-23669-01

03029711

Dear Mr. Rolph,

This letter is to delete a temporary storage site from our NRC license located at:

4606 Bainbridge Boulevard
Chesapeake, VA 23320

Branscome has never had a nuclear density gauge leak test to show positive. Attached are copies of the last leak test from our gauges that may have been stored at this location.

Also please add temporary storage sites at:

Branscome Inc.
Dominion Asphalt Plant
120 South Dominion Boulevard
Chesapeake, VA 23320

And

NIT
Branscome Construction Trailer
7737 Hampton Boulevard
Norfolk, VA 23505

Any questions please contact me at my office: 757-247-1354, or mobil: 757-342-8995.

Sincerely yours,

Michael C. Dudley
Quality Control Manager
Branscome Incorporated

cc: W. Jay Lipscomb, Branscome Inc.

135777

NMSS/RGNI MATERIALS-002





Troxler Electronic Laboratories, Inc.

3008 Cornwallis Rd., P.O. Box 12057
 Research Triangle Park, NC 27709
 Tel: (877) 876-9537 Fax: (919) 485-2250

License: NC 032-0182-1

LEAK TEST CERTIFICATE

DEVICE:

Model: 4640 Serial No: 855

SEALED SOURCES:

SERIAL NO.	MEASURE DATE	NUCLIDE	ACTIVITY	
			(GBq)	(mCi)
50-8002	1/29/1990	CS-137	0.296	8

LEAK TEST ANALYSIS:

Sample collected on: 09/24/2004

Sample analyzed on: 10/04/2004 at 1:17:00 PM

Analyzed by: J. Oisen

	ALPHA	BETA-GAMMA
Conversion factor (cpm/Bq)	1.30E+01	2.11E+01
Background measurement (cpm)	1	29
Sample measurement (cpm)	0	30
Activity (Bq)	< MDA	< MDA
Min. Detectable Activity (Bq)	5.7E-01	1.3E+00

This certifies that the above leak test results are:

Less than 185 Bq (0.005 uCi)
 Greater than 185 Bq (0.005 uCi)

If greater than 185 Bq (0.005 uCi):

Person Notified _____ Date _____
 Phone _____ and/or Fax _____



Troxler Electronic Laboratories, Inc.

3008 Cornwallis Rd., P.O. Box 12057
Research Triangle Park, NC 27709
Tel: (877) 876-9537 Fax: (919) 485-2250

License: NC 032-0182-1

LEAK TEST CERTIFICATE

DEVICE:

Model: 4640 Serial No: 680

SEALED SOURCES:

SERIAL NO.	MEASURE DATE	NUCLIDE	ACTIVITY	
			(GBq)	(mCi)
50-6868	6/29/1989	CS-137	0.296	8

LEAK TEST ANALYSIS:

Sample collected on: 09/24/2004

Sample analyzed on: 09/30/2004 at 11:09:00 AM

Analyzed by: J. Olsen

	ALPHA	BETA-GAMMA
Conversion factor (cpm/Bq)	1.30E+01	2.11E+01
Background measurement (cpm)	1	26
Sample measurement (cpm)	1	18
Activity (Bq)	< MDA	< MDA
Min. Detectable Activity (Bq)	5.7E-01	1.3E+00

This certifies that the above leak test results are:

Less than 185 Bq (0.005 uCi)

Greater than 185 Bq (0.005 uCi)

If greater than 185 Bq (0.005 uCi):

Person Notified _____ Date _____

Phone _____ and/or Fax _____



Troxler Electronic Laboratories, Inc.

3008 Cornwallis Rd., P.O. Box 12057
 Research Triangle Park, NC 27709
 Tel: (877) 876-9537 Fax: (919) 485-2250

License: NC 032-0182-1

LEAK TEST CERTIFICATE

DEVICE:

Model: 4640 Serial No: 400

SEALED SOURCES:

SERIAL NO.	MEASURE DATE	NUCLIDE	ACTIVITY	
			(GBq)	(mCi)
50-4427	1/22/1988	CS-137	0.296	8

LEAK TEST ANALYSIS:

Sample collected on: 09/24/2004

Sample analyzed on: 10/04/2004 at 1:26:00 PM

Analyzed by: J. Olsen

	ALPHA	BETA-GAMMA
Conversion factor (cpm/Bq)	1.30E+01	2.11E+01
Background measurement (cpm)	1	29
Sample measurement (cpm)	0	26
Activity (Bq)	< MDA	< MDA
Min. Detectable Activity (Bq)	5.7E-01	1.3E+00

This certifies that the above leak test results are:

- Less than 185 Bq (0.005 uCi) Greater than 185 Bq (0.005 uCi)

If greater than 185 Bq (0.005 uCi):

Person Notified _____ Date _____

Phone _____ and/or Fax _____



Troxler Electronic Laboratories, Inc.

3008 Cornwallis Rd., P.O. Box 12057
 Research Triangle Park, NC 27709
 Tel: (877) 876-9537 Fax: (919) 485-2250

License: NC 032-0182-1

LEAK TEST CERTIFICATE

DEVICE:

Model: 4640 Serial No: 384

SEALED SOURCES:

SERIAL NO.	MEASURE DATE	NUCLIDE	ACTIVITY	
			(GBq)	(mCi)
50-4331	12/9/1987	CS-137	0.296	8

LEAK TEST ANALYSIS:

Sample collected on: 09/24/2004

Sample analyzed on: 09/30/2004 at 11:08:00 AM

Analyzed by: J. Olsen

	ALPHA	BETA-GAMMA
Conversion factor (cpm/Bq)	1.30E+01	2.11E+01
Background measurement (cpm)	1	26
Sample measurement (cpm)	0	32
Activity (Bq)	< MDA	< MDA
Min. Detectable Activity (Bq)	5.7E-01	1.3E+00

This certifies that the above leak test results are:

Less than 185 Bq (0.005 uCi) Greater than 185 Bq (0.005 uCi)

If greater than 185 Bq (0.005 uCi):

Person Notified _____ Date _____

Phone _____ and/or Fax _____



Troxler Electronic Laboratories, Inc.

3008 Cornwallis Rd., P.O. Box 12057
 Research Triangle Park, NC 27709
 Tel: (877) 876-9537 Fax: (919) 485-2250

License: NC 032-0182-1

LEAK TEST CERTIFICATE

DEVICE:

Model: 3450 Serial No: 599

SEALED SOURCES:

SERIAL NO.	MEASURE DATE	NUCLIDE	ACTIVITY	
			(GBq)	(mCi)
47-28711	3/10/1999	AM-241:BE	1.48	40
750-7883	1/16/2001	CS-137	0.296	8

LEAK TEST ANALYSIS:

Sample collected on: 09/24/2004

Sample analyzed on: 09/30/2004 at 11:07:00 AM

Analyzed by: J. Olsen

	ALPHA	BETA-GAMMA
Conversion factor (cpm/Bq)	1.30E+01	2.11E+01
Background measurement (cpm)	1	28
Sample measurement (cpm)	0	24
Activity (Bq)	< MDA	< MDA
Min. Detectable Activity (Bq)	5.7E-01	1.3E+00

This certifies that the above leak test results are:

- Less than 185 Bq (0.005 uCi) Greater than 185 Bq (0.005 uCi)

If greater than 185 Bq (0.005 uCi):

Person Notified _____ Date _____

Phone _____ and/or Fax _____



Troxler Electronic Laboratories, Inc.

3008 Cornwallis Rd., P.O. Box 12057
 Research Triangle Park, NC 27709
 Tel: (877) 876-9537 Fax: (919) 485-2250

License: NC 032-0182-1

LEAK TEST CERTIFICATE

DEVICE:

Model: 3411 Serial No: 8996

SEALED SOURCES:

SERIAL NO.	MEASURE DATE	NUCLIDE	ACTIVITY	
			(GBq)	(mCi)
40-6301	3/17/1982	CS-137	0.296	8
47-5464	3/4/1982	AM-241:BE	1.48	40

LEAK TEST ANALYSIS:

Sample collected on: 09/24/2004

Sample analyzed on: 10/04/2004 at 1:16:00 PM

Analyzed by: J. Olsen

	ALPHA	BETA-GAMMA
Conversion factor (cpm/Bq)	1.30E+01	2.11E+01
Background measurement (cpm)	1	29
Sample measurement (cpm)	0	20
Activity (Bq)	< MDA	< MDA
Min. Detectable Activity (Bq)	5.7E-01	1.3E+00

This certifies that the above leak test results are:

Less than 185 Bq (0.005 uCi)

Greater than 185 Bq (0.005 uCi)

If greater than 185 Bq (0.005 uCi):

Person Notified _____ Date _____

Phone _____ and/or Fax _____