

From: "Fassell, John (DHS-DFDRS-RHB-RMS)" <JFassell@dhs.ca.gov>
To: "Sandra Gabriel" <slg2@nrc.gov>
Date: Thu, Jan 20, 2005 4:08 PM
Subject: RE: Alpha-Omega's Authorization to Install Nucletron ClassicHDR

The linear accelerator would constitute an existing irradiation facility.

-----Original Message-----

From: Sandra Gabriel [mailto:slg2@nrc.gov]
Sent: Thursday, January 20, 2005 12:45 PM
To: Fassell, John (DHS-DFDRS-RHB-RMS)
Subject: Re: Alpha-Omega's Authorization to Install Nucletron ClassicHDR

This situation is not installation at an existing HDR irradiation facility--this is the initial installation of a used unit at the facility of the new purchaser. Please confirm that this is authorized. Thank you.

(there is also a linear accelerator at this location, if you would consider this to constitute an "existing irradiation facility")

Sandy Gabriel
Senior Health Physicist
Medical Branch
NRC Region I

>>> "Fassell, John (DHS-DFDRS-RHB-RMS)" <JFassell@dhs.ca.gov> 01/20/05 15:29 PM >>>
Alpha-Omega's license (#3925-27) currently reads for License Condition 9EE "source exchange and maintenance" on the Nucletron-Odeft MicroSelectron HDR Afterloader. The license reviewer that did the amendment understood maintenance to include installation at existing irradiation equipment installations. I realize that the verbage is different for this device than for some of the other things that they are allowed to do but the intent was to include installation.

John Fassell
Associate Health Physicist
e-mail:jfassell@dhs.ca.gov
916-440-7923/Fax:916-341-7103

Mail Envelope Properties

(41F01DCD.F45 : 19 : 48965)

Subject: RE: Alpha-Omega's Authorization to Install Nucletron ClassicHDR
Creation Date: Thu, Jan 20, 2005 4:08 PM
From: "Fassell, John (DHS-DFDRS-RHB-RMS)" <JFassell@dhs.ca.gov>
Created By: JFassell@dhs.ca.gov

Recipients

nrc.gov
kp1_po.KP_DO
SLG2 (Sandra Gabriel)

Post Office
kp1_po.KP_DO

Route
nrc.gov

Files	Size	Date & Time
MESSAGE	1374	Thursday, January 20, 2005 4:08 PM
Mime.822	2690	

Options

Expiration Date: None
Priority: Standard
Reply Requested: No
Return Notification: None

Concealed Subject: No
Security: Standard