

U.S. NUCLEAR REGULATORY COMMISSION		Conversation Date: 1/24/2005	
TELEPHONE CONVERSATION RECORD		Time: 12:40 pm	
Mail Control No.:	License No.:	Docket No.:	
136309	37-00993-05	030-08702	
Licensee/Applicant Participant(s):	Organization:	Telephone No.:	
John Babu (RSO)	Mercy Health of Philadelphia	610-237-4296	
Person(s) Calling: Joe Nick			
Subject: Request for Additional Information			
<p>Summary:</p> <p>I spoke to the licensee's RSO on 1/24/05 and asked for additional information regarding the recent letter (dated 1/19/05) requesting an expedited amendment for the NRC material license.</p> <p>The licensee's letter requested the addition of an area for the storage of radioactive waste generated from the administration of materials listed in 10 CFR 35.300. The licensee's request contained a sketch showing the proposed storage area in the sub-basement of the facility. The request also stated that the area was controlled, away from main traffic, and had walls with at least 8 inches of concrete. There was no information regarding what was above or below the proposed storage area.</p> <p>The RSO stated that there was nothing below the storage area, and that a ground-floor hallway was above the room. In addition, the RSO stated that the ceiling was composed of 5 to 6 inches of concrete.</p> <p>I told the RSO that we had enough information to issue an amendment.</p> <p>The fax number for the RSO is 610-237-5718.</p>			
Action Required/Taken: Issue license amendment for the new waste storage area.			
Prepared By: Joe Nick		Date: 1/24/2005	