

U.S. NUCLEAR REGULATORY COMMISSION		Conversation Date: 01/14+24/05	
TELEPHONE CONVERSATION RECORD		Time: various	
Mail Control No.: 136170	License No.: 29-00232-08	Docket No.: 03002423	
Licensee/Applicant Participant(s): Karen Wheeler, physics consultant	Organization: St. Francis Med Ctr	Telephone No.: 908-788-9440, ext. 905	
Person(s) Calling: SGabriel, RI			
Subject: Amendment request			
<p>Summary:</p> <p>Please provide:</p> <ul style="list-style-type: none"> a) room diagram b) description of shields c) procedure to assure shields will remain in intended position to reduce rad. levels in surrounding uncontrolled areas to <2 mR/hr <p>Fax to 610-337-5269, referencing mail control 136170.</p>			
Action Required/Taken: Await response from licensee			
Prepared By: SGabriel		Date: 1/25/05	