

U.S. NUCLEAR REGULATORY COMMISSION		Conversation Date: 1/26/05	
TELEPHONE CONVERSATION RECORD		Time: 10:30am	
Mail Control No.:	License No.:	Docket No.:	
136300	37-30073-01	03033294	
Licensee/Applicant Participant(s):	Organization:	Telephone No.:	
Jeff Linder, Director of Hospital and Patient Ser.	American Red Cross	570-821-2406	
Person(s) Calling: J. A. Joustra			
Subject: Letter dated 1/17/05			
Summary:			
<ol style="list-style-type: none"> 1. Need management signature. Mr Linder is management rep. 2. Need a step by step procedure for analyzing leak tests. Since Ms Anderko is not a Service licensee. The procedure will be placed in the American Red Cross license file. We need specifics on the instrument used, method for calculating Eff. And MDA. 			
Action Required/Taken: Wait for additional info and review and amend the license if the info is acceptable.			
Prepared By: J. Joustra		Date: 1/26/05	