147 Gettys Street P.O. Box 3786 Gettysburg, PA 17325-0786 717.334.2121 Tel www.wellspan.org





*05 JAN 19 A7:20

January 11, 2005

U.S. Nuclear Regulatory Commission Materials Licensing Branch Region I 475 Allendale Road King of Prussia, PA 19406-1415

RE:

Closeout Survey

Radioactive Material License No. 37-16826-01

Gettysburg Hospital

03011706

Gentlemen:

We have decommissioned our old nuclear medicine area. All radioactive materials have been removed from the area. The following is a closeout survey performed on January 3, 2005.

Gross contamination smears were taken on all walls and on the floor and doors. The smears were counted in a Capintec Captus 2000 well counter with an efficiency for Tc-99m which is greater than 90% and minimum detectable activity of 102 dpm. Area radiation level surveys were taken at each smear site. The background for the area surveys was 0.02 mR/hr using a Ludlum 14C GM survey meter last calibrated on June 15, 2004. All readings were 0.02 mR/hr.

The smear results for the area are contained on the following pages. The results indicate that the area is free of contamination.

All radioactive material has been removed from the site and we have released the room for unrestricted use.

Sincerely,

Joseph Edgar, MHA

Senior Vice President for Operations

Enclosures

136317

NMGG/RGNI MATERIALS-002

JE/mz

Gettysburg Hospital - Wellspan Imaging Services- Nuclear Medicine Vertex Classic Scan Room - Final Survey

January 3, 2005

The following are the results of the final survey of the old Vertex Classic scan room:

Area Survey using L14C-2 GM survey meter: 0.02 mR/hr BKG

0.02 mR/hr all surfaces in room

WIPE SURVEYS
Background = 364 cpm

Area Surveyed	Measured Activity
entrance doorway -outside	20 dpm
entrance doorway- inside	20 dpm
entrance door handles	20 dpm
entrance door frame	20 dpm
floor at entrance doorway	20 dpm
wall (right)	51 dpm
back wall	51 dpm
wall (left)	51 dpm
front wall	51 dpm
office door frame	37 dpm
office door handles	37 dpm
floor at office doorway	37 dpm
central/general floor area	60 dpm

Gettysburg Hospital - Wellspan Imaging Services- Nuclear Medicine Hot Lab - Final Survey

January 3, 2005

The following are the results of the final survey of the old Nuclear Medicine Hot lab:
Area Survey using L14C-2 GM survey meter: 0.02 mR/hr BKG
0.02 mR/hr all surfaces in room

WIPE SURVEYS
Background = 364 cpm

Area Surveyed	Measured Activity
doorway -outside	22 dpm
doorway- inside	22 dpm
door handles	22 dpm
door frame	22 dpm
floor at doorway	22 dpm
front wall	45 dpm
lead window	45 dpm
light switch	45 dpm
sink area/wall	0 dpm
sink handles	0 dpm
sink	0 dpm
drain	0 dpm
stainless steel (s.s.) counter	0 dpm
wall above s.s. counter	0 dpm
cabinets below counter	0 dpm
floor beneath s.s. counter	0 dpm
Xenon hood	29 dpm
left side work area-walls	15 dpm
work area counter tops	15 dpm
cabinets above work area	15 dpm
cabinets below work area	15 dpm
floor beneath work area	15 dpm
center floor area	32 dpm

Gettysburg Hospital - Wellspan Imaging Services- Nuclear Medicine ARC Scan Room - Final Survey

January 3, 2005

The following are the results of the final survey of the old ARC scan room:

Area Survey using L14C-2 GM survey meter: 0.02 mR/hr BKG 0.02 mR/hr all surfaces in room

WIPE SURVEYS
Background = 364 cpm

Area Surveyed	Measured Activity
doorway -outside	16 dpm
doorway- inside	16 dpm
door handles	16 dpm
door frame	16 dpm
floor at doorway	16 dpm
wall (right)	32 dpm
back wall	32 dpm
wall (left)	32 dpm
front wall	32 dpm
work counter	0 dpm
cabinets above counter	0 dpm
cabinets beneath counter	0 dpm
floor beneath counter	26 dpm
central/general floor area	26 dpm

Gettysburg Hospital - Wellspan Imaging Services- Nuclear Medicine Vertex Plus Scan Room - Final Survey

January 3, 2005

The following are the results of the final survey of the old Vertex Plus scan room:

Area Survey using L14C-2 GM survey meter: 0.02 mR/hr BKG 0.02 mR/hr all surfaces in room

WIPE SURVEYS
Background = 364 cpm

Area Surveyed	Measured Activity
entrance doorway -outside	53 dpm
entrance doorway- inside	53 dpm
entrance door handles	53 dpm
entrance door frame	53 dpm
floor at entrance doorway	53 dpm
wall (right)	11 dpm
window sill	11 dpm
back wall	11 dpm
wall (left)	11 dpm
front wall	11 dpm
sink	26 dpm
central/general floor area	0 dpm

Gettysburg Hospital - Wellspan Imaging Services- Nuclear Medicine Vertex Classic Scan Room - Final Survey

January 3, 2005

The following are the results of the final survey of the old Vertex Classic scan room:

Area Survey using L14C-2 GM survey meter: 0.02 mR/hr BKG

0.02 mR/hr all surfaces in room

WIPE SURVEYS
Background = 364 cpm

Area Surveyed	Measured Activity
entrance doorway -outside	20 dpm
entrance doorway- inside	20 dpm
entrance door handles	20 dpm
entrance door frame	20 dpm
floor at entrance doorway	20 dpm
wall (right)	51 dpm
back wall	51 dpm
wall (left)	51 dpm
front wall	51 dpm
office door frame	37 dpm
office door handles	37 dpm
floor at office doorway	37 dpm
central/general floor area	60 dpm

• .	ne receipt of your letter/application dated	:
1/11/2005 ncludes an administration	and to inform you that the initial processing which review has been performed.	:
1 some 3	7-16826-01	•
There were no admin technical reviewer. P	istrative omissions. Your application was assigned to a lease note that the technical review may identify additional additional information.	
Please provide to this	office within 30 days of your receipt of this card	
Please provide to this	office within 30 days of your receipt of this card	
Please provide to this	office within 30 days of your receipt of this card	
A copy of your action ha	office within 30 days of your receipt of this card s been forwarded to our License Fee & Accounts Receivable you separately if there is a fee issue involved.	
A copy of your action ha Branch, who will contact Your action has been as When calling to inquire a	s been forwarded to our License Fee & Accounts Receivable	
A copy of your action ha Branch, who will contact Your action has been as When calling to inquire a	s been forwarded to our License Fee & Accounts Receivable you separately if there is a fee issue involved. signed Mail Control Number	

.

,

BETWEEN:		: (FOR LFMS USE) : INFORMATION FROM LTS :
License Fee Management Br and Regional Licensing Sectio		Program Code: 02120 Status Code: 0 Fee Category: 7C Exp. Date: 20130831 Fee Comments: Decom Fin Assur Reqd: N
LICENSE FEE TRANSMITTAL		
A. REGION I		
License No.:		PITAL WELLSPAN
2. FEE ATTACHED Amount: Check No.:		
3. COMMENTS	Signed _ Date _	M. a. Lection
B. LICENSE FEE MANAGEMENT	BRANCH (Check	when milestone 03 is entered //)
1. Fee Category and Amou	nt:	
2. Correct Fee Paid. Ap Amendment Renewal License	plication may l	be processed for:
3. OTHER	Signed	

Date