



Main Line Cardiology Assoc., P.C.

Primary, Nuclear and Interventional Cardiology

RECEIVED
REGION 1

Elliot M. Gerber, M.D., F.A.C.C.
Antoinette M. Sulpizi, M.D., F.A.C.C.
Neal F. Skop, M.D., F.A.C.C.
Andrew R. Bowman, M.D., F.A.C.C.
Nirav N. Mehta, M.D.
Renee Smith, PA-C
Dawn Dwyer, PA-C

Paoli Memorial Medical Bldg. #2, Suite 328
Paoli, PA 19301
Telephone (610) 647-2400
Fax (610) 647-3902

'05 JAN 24 P12:45

MS16

K-4

37-30979-01

03036774

January 17, 2005

Steven Courtemanche
US NRC Region I
475 Allendale Rd.
King of Prussia, PA 19406-1415

Re : Request for additional information, Re: Control Number 136107.

Dear Mr. Courtemanche,

Please accept the attached supplemental information per your request.

1. Contact information for our Radiation Safety Officer.

Dr. Bowman is a member of our practice and may be reached through our office at any time at the following numbers:

Phone: (610) 647-2400

Fax: (610) 647-3902

2. Location of Use.

We verify, to the best of our knowledge, that our proposed location of use is not a location of use specified on any other active NRC license.

3. Use of Co-57.

We understand the licensing requirements of Co-57 and have already obtained a Pennsylvania State radioactive materials license for its use.

4. Controlled vs. Restricted Areas.

According to the strict definitions of these terms found in 10 CFR 20.1003, the hot lab, camera room and the stress room will be considered as restricted areas. These rooms will normally be accessed only by facility staff and those with a need to be there, i.e., patients, physicist, those providing patient care, etc. The other rooms in the facility will be considered as controlled areas, access to which may be limited as needed.



136107

NMSS/RGNI MATERIALS-002

5. Calibration of Equipment used to Measure Patient Doses.

We confirm that we will use nationally recognized standards when calibrating and performing routine quality control on our radiation detection equipment that is used to measure patient doses.

6. Written Procedures.

Written procedures will be maintained in a Radiation Protection Plan and will be available for review by the NRC during any inspection of our facility.

We value your input to our program. If you have any questions or need any additional information regarding this application, please feel free to contact me at your convenience at (610) 647-2400.

Sincerely,

A handwritten signature in black ink, appearing to read 'Elliot M. Gerber', with a stylized flourish at the end.

Elliot M. Gerber, M.D.
Managing Partner