Odyssey Imaging "The Future Of Healthcare"
P.O. Box 4393
Roanoke, VA 24015

December 30,2004

P-7

USNRC Region I
Materials Licensing Section
Mr. Tom Thompson
475 Allandale Rd
King of Prussia, PA 19406

RE: License # 45-25516-01 03035377

Dear Mr. Thompson:

I am writing to inform you that the Nuclear Medicine Lab previously located at 1935 W Main St, Salem, VA has been moved to Primary Care Associates, PC 1955 W. Main St, Salem, VA as stated in our last amendment change to you. The old area has been surveyed and wipe tested. All equipment, radioactive materials and records are now relocated to the new area. All other conditions remain the same.

Please find below the final surveys and wipe tests for the old area. You may wish to delete the old address from your records.

Decommissioned camera room and hot lab 11/18/2004

	Surveys	Wipes
Hot lab dose prep counter	0.03 mr/hr	118cpm
Hot lab dose prep floor	0.03 mr/hr	65cpm
Waste storage	0.03 mr/hr	109cpm
Camera Area Floor	0.03mr/hr	84cpm
Processing Area	0.03mr/hr	133cpm
Injection Area	0.03mr/hr	112cpm
Treadmill	0.03mr/hr	139cpm
Background	0.03mr/hr	118cpm

Survey Meter Used—Ludlum model 14C serial #201570, Cal date 1/24/2004
Scaler Ratemeter Used—Ludlum Model 220, Serial #198605, Cal date 1/09/2004

If you have any questions or concerns please do not hesitate to call me at my office (540) 563-9840.

Thank you,

Martin Hellkamp, President
Odyssey Imaging

136230
NMSS/RGNI MATERIALS-002