

**From:** Thomas Thompson  
**To:** Internet:Englishl@dickinson.edu  
**Date:** Thu, Feb 10, 2005 11:26 AM  
**Subject:** Your January 25, 2005 letter regarding NRC license renewal.  
**Place:** LAT@nrc.gov

37-09726-02(03000994)

Thank you for your letter. Some additional details are needed with regard to instrument calibration. As indicated in my January 20th e-mail NRC's NUREG-1556, Vol. 7 provides acceptable procedures in Appendix M. If you do not want to subscribe to these procedures you should provide procedures with equivalent detail.

You may reply by facsimile to 610 337 5269 and refer to Mail Control 136200.

Thank you again.

**Mail Envelope Properties**

(420B8B22.999 : 23 : 35376)

**Subject:** Your January 25, 2005 letter regarding NRC license renewal.

**Creation Date:** Thu, Feb 10, 2005 11:26 AM

**From:** Thomas Thompson

**Created By:** TKT@nrc.gov

**Recipients**

dickinson.edu

englishl (Internet:Englishl@dickinson.edu)

**Post Office****Route**

dickinson.edu

**Files**

MESSAGE

**Size**

1098

**Date & Time**

Thursday, February 10, 2005 11:26 AM

**Options**

**Expiration Date:** None

**Priority:** Standard

**Reply Requested:** No

**Return Notification:** None

**Concealed Subject:** No

**Security:** Standard