

ATTACHMENT 7

RESULTS OF ACTIVITY DETECTED DURING SURFACE SCANS

Surface Activity Measurements of Residual Contamination Identified
During Surface Scans

Area Designation	Total Surface Activity in DPM/100cm ²		
	Gross Alpha	Uncertainty ¹	MDA
5462B - wall B-1 (C6)	28	12	21
5462B - wall B-2 (C6)	30	12	21
5462B - wall C-3 (D7, E7)	24	12	21
5462B - wall C-4 (D7, E7)	35	13	21
5462B - wall C-5 (D7, E7)	23	12	21
3830 - Benchtopy (H4)	33	13	22
3830 - Benchtopy (H4)	26	12	22
3830 - Floor (I6)	29	13	22

¹ - at the 95% confidence level

Surface Activity Measurements of Residual Contamination Identified
During Surface Scans

Area Designation	Total Surface Activity in DPM/100cm ²		
	Gross Beta	Uncertainty ¹	MDA
6760A-Casework, base- J6	3660	329	413
6046A - benchtop- G6:H6 (verification survey)	1595	281	406
5880, casework, C1 - (E6, F6)	3337	320	408
5880, casework, C1 - (E6, F6)	4120	336	408
5880, casework, C1 - (E6, F6)	3419	322	408
5880, casework, C1 - (E6, F6)	605	257	408
5880, casework, C1 - (E6, F6)	1457	278	408
5880, casework, C1 - (E6, F6)	1289	274	408
5880, casework, C1 - (E6, F6)	1632	282	408
5880, casework, C1 - (E6, F6)	2144	294	408
5880, casework, C1 - (E6, F6)	2488	302	408
5880, casework, C1 - (E6, F6)	1014	268	408
5880, casework, C1 - (E6, F6)	911	265	408
5880, casework, C1 - (E6, F6)	3333	320	408
5880, casework, C1 - (E6, F6)	3251	318	408
5880, casework, C1 - (E6, F6)	3148	316	408
5880, casework, C1 - (E6, F6)	3450	322	408
5880, casework, C1 - (E6, F6)	3323	320	408
5880, casework, C1 - (E6, F6)	2701	306	408
5880, casework, C1 - (E6, F6)	2086	293	408
3838A-Stainless steel benchtop, cup sink - (E6)	2392	309	427
3838A- Floor (G6, H6)	3430	472	709
3838A- Floor (G6, H6)	5120	495	709
3838A-Stainless steel benchtop, -(E6)	1914	296	421
3838A- Stainless steel benchtop, vacuum nozzle (E6)	3223	323	421

¹ - at the 95% confidence level