



Sentara Virginia Beach General Hospital
1080 First Colonial Road
Virginia Beach, Virginia 23454

www.sentara.com

February 11, 2005

US Nuclear Regulatory Commission
Region II
Atlanta Federal Center
51 Forsyth Street, S.W., Suite 23T85
Atlanta, Georgia 30303-3415

J-9

Dear Ms. Weidner,

03003348

At this time, we request the addition of Robert Mariano, M.D., as an authorized user to NRC license #45-11035-01. We request that he be added to the above stated license and authorized for medical use as described in 10CFR 35.100 and 35.200.

For documentation of training and experience, please refer to NRC license # 45-00986-01. Dr. Mariano is currently authorized for the above medical use categories under that hospital based license issued to Bon Secours - DePaul Medical Center.

The following information is pursuant to our telephone conversation on 2/7/2005. In order to remove the "in vitro" category from our license, you requested wipes and surveys to be performed from the areas in which the tests were performed. I have performed six surveys and six wipes in areas representative of use, storage, and disposal activities.

The results are as follows:

| | Wipes (DPM) | Survey (mr/hr) |
|--------------------------------|-------------|----------------|
| Refrigerator 1 | < Bkg | .02 |
| Refrigerator 2 | < Bkg | .02 |
| Storage cabinet | < Bkg | .02 |
| Floor adjacent to cabinet | < Bkg | .02 |
| Work counter | < Bkg | .02 |
| Floor adjacent to work counter | < Bkg | .02 |
| Background | 0 | .02 |

I hope this information fulfills your needs for approval of this request and the prior request already in your possession. As always, do not hesitate to contact me if you require any further information. I can be reached at (757) 395-8920 or by e-mail.

Respectfully,

Blair Mazzocco

Blair Mazzocco, CNMT
RSO

136104

NMSS/RGNI MATERIALS-002