UNITED STATES
NUCLEAR REGULATORY COMMISSION
REGION IV
611 RYAN PLAZA DRIVE, SUITE 400
ARLINGTON, TEXAS 76011-8064

F A C S I M I L E

Name:	Gary L. Dodson	
Organization:	City of Cheyenne	License No. 49-18179-01 Docket No. 030-14612 Control No. 470430
Fax:	307-637-6454	
Phone:	307-637-6263	
From:	Jacqueline D. Cook	
Date:	March 14, 2005	
Subject:	Renewal Application dated February 15, 2005	
Pages:	2 (8 with Appendix B included)	

Mr. Dodson:

Per your renewal application dated February 15, 2005, and your fax response dated March 11, 2005, the items on the next page are additional deficiencies which require your response. Please respond to this fax by Monday, March 28, 2005. Our fax number is (817) 860-8263. If you have any questions regarding this fax, please call me at (817) 860-8132. When responding to this fax, please include the license, docket, and control numbers located at the top of this page.

Thanking you in advance for your cooperation and assistance in this matter.

/RA/
Jacqueline D. Cook
Senior Health Physicist

1. Item 10, *"Radiation Safety Program - Operating and Emergency Procedures"* of Appendix B, page B-4, you checked both options.

   Please commit to **one** of the following:

   , "We will implement and maintain the operating ane emergency procedures in Appendix H of NUREG-1556, Vol. 1, Rev. 1, 'Consolidated Guidance about Materials Licenses: Program-Specific Guidance about Portable Gauge Licenses,' dated November 2001, and provide copies of these procedures to all gauge users and at each job site."

   OR

   , "Operating and emergency procedures will be developed, implemented, and maintained, and will meet the criteria in the section entitled 'Radiation Safety Program - Operating and Emergency Procedures' in NUREG-1556, Vol. 1, Rev. 1, 'Consolidated Guidance about Materials Licenses: Program-Specific Guidance about Portable Gauge Licenses,' dated November 2001."

2. Item 10, *"Radiation Safety Program - Maintenance"* of Appendix B, page B-5, you inadvertently omitted a response to **non-routine maintenance or repair operations that require detaching the source or source rod from the gauge.**

   Please submit either of the following:

   , A statement that: "We will send the gauge to the manufacturer or other person authorized by NRC or an Agreement State to perform non-routine maintenance or repair operations that require detaching the source or source rod from the gauge."

   OR

   , The information listed in Appendix G supporting a request to perform this work in-house.