

Pfizer Inc.  
7000 Portage Road  
Kalamazoo, MI 49001

---



Dee L. Clement  
Senior Advisor – RSO  
tel 269-833-9431  
fax 269-833-9400  
dee.l.clement@pfizer.com

March 8, 2005

Dr. Peter Lee  
Nuclear Regulatory Commission, Region III  
2443 Warrenville Road STE 210  
Lisle, IL 60532-4352

Subject: NRC License #21-00182-03 amendment #62 - Jasper Clinic Decommissioning Information

Dear Dr Lee:

This letter is in response to our phone conversation today concerning the results of the pancake probe surveys for the Jasper Clinic (Building 258) decommissioning. Pharmacia and Upjohn Company performed scans on the surfaces of the impacted rooms using the GSM 110 + pancake probe identified on the "Decommissioning Form" and no contamination was detected in the rooms at the Jasper Clinic facility, all results were below 50 ncpm. The results from these pancake probe surveys were inadvertently left out of the decommissioning documentation so please add this note to the Jasper Clinic Decommissioning Enclosure 1-B in the amendment #62 packet. Contact me if you have any questions.

Sincerely,

A handwritten signature in cursive script that reads "Dee L. Clement".

Dee L. Clement  
Radiation Safety Officer - Kalamazoo Site

MAR 14 2005



PRGD - DUCLEMENT (30841)  
Pfizer Inc  
301 Henrietta  
Kalamazoo, MI 49007

POSTAGE  
PAID



309

U.S. POSTAGE

00.309

METER 701613

**Global Research & Development**

TO: DR. PETER LEE  
USNRC - REGION III  
2443 WARRENVILLE Rd STE 210  
LISLE, IL 60532-4352

60532

