

CONVERSATION RECORD

TYPE:

Outgoing Telephone XXX
Incoming Telephone
Meeting

NAME OF PERSON CONTACTED:
John Miller, RSO

ORGANIZATION:
International Isotopes
Inc.

TIME:
3:06pm

DATE:
03/8/05

SUBJECT: Question regarding International Isotopes Inc. response to the NRC's request for additional information regarding their application for registration of irradiator and teletherapy sources.

SUMMARY: Mr. Miller, was contacted to determine the reason for the changes to the minimum lengths of the sources on the new engineering drawings submitted with the response to the request for additional information. He stated that they had decided to make the minimum lengths of the capsules smaller based on a request from potential customers once the sources have been approved.

ACTION REQUIRED: Additional information to be included in the registration certificate regarding the minimum lengths of the sources.

PLACE THIS RECORD IN:

Registration File XXX
QA File
Incident File
General File

**PERSON DOCUMENTING THE
CONVERSATION:**
Tomas Herrera

SIGNATURE:
/RA/

DATE:
03/09/05

--	--	--