

TELEPHONE CONVERSATION RECORD

| | | |
|---|-----------------------------------|----------------------------|
| Licensee Name: P&G -Clairol, Inc. | Date: 8/4/2004 | Time: 2 PM |
| License No.: 06-11703-02 | Docket No.: 03003829 | Control No.: 135260 |
| Licensee Contact: David Erlingheuser, RSO | Telephone No: 203.357.5045 | |
| NRC Contact: Kathy Dolce Modes | Telephone No: 610.337.5251 | |
| Subject: final status survey and license termination | | |
| Summary: <ol style="list-style-type: none">1. Prior to termination, 10 CFR 30.36(l)(4) requires that you submit to the NRC all records required by 10 CFR 30.51(d and f) and 30.35(g).2. Please provide the square footage of the entire building.3. Provide a description of the Blachey Road facility [such as "residential", "industrial", "commercial", "mixed residential/commercial", etc.].4. Survey records only discuss H-3 and C-14. How did you survey for S-35 (87 day half-life)?5. In a letter dated June 23, 1994, Clairol provided a copy of a floor diagram for the basement and first floor of the Blachley Road facility. Room BN-11 ("Pack Lab") was highlighted and Instrument Room IN-21 on the first floor was also highlighted. In your May 17, 1995 letter you noted that the "Counting Room" was no longer used for counting samples in the liquid scintillation counter. Please explain if radioactive material had ever been used and/or stored in these laboratories. If material was in use or in storage, please provide the final status surveys for these rooms. If the room numbers or room nomenclature has changed, please provide clarification. As you know, we will need to ensure that all areas where radioactive material had been used and/or stored are appropriately surveyed. | | |
| Action Required/Taken: Wait for additional information from licensee | | |
| Signature: Kathy Dolce Modes | Date: 8/4/2004 | |