

Nicholas Smarra
Medical Physicist
2 Crescent Park West
Warren, PA 16365

Sandra Gabriel
Senior Health Physicist
Medical Branch
NRC Region I

MS-16
K-8

March 8, 2005
License # 37-17863-01 03013549

Ms. Gabriel

I am writing in response to the following question that you had regarding the release criteria for Cs-131

I have reviewed your response. Regarding release criteria for Cs-131 patients, am I correct in concluding that your CF is 83.8, therefore you will release patients with dose rate at 1 meter of no more than $(500/83.8)=6$ mR/hr?

This is a correct assumption. I apologize for not filing that area in when sending you the criteria. I hope this clears everything up and we will be looking for the amendment shortly.

Sincerely,



Nicholas J Smarra
Medical Physicist

136247
NMSS/RGNI MATERIALS-002