

U.S. NUCLEAR REGULATORY COMMISSION		Conversation Date: 3/28/05	
TELEPHONE CONVERSATION RECORD		Time: 1:00 p.m.	
Mail Control No.:	License No.:	Docket No.:	
135837	37-30085-01	03033332	
Licensee/Applicant Participant(s):	Organization:	Telephone No.:	
Paul Castro, Ph.D.	Jefferson Radiation Oncology Center	724-422-3310	
Person(s) Calling: SGabriel, RI			
Subject: License renewal			
<p>Summary:</p> <p>We received your deficiency response faxed on March 8, 2005. Please confirm that the following will be included in your spot-checks for compliance with 10 CFR 35.643(d):</p> <ul style="list-style-type: none"> a) verification of clock (date and time) in HDR unit's computer b) verification of decayed source activity in unit's computer <p>Paul confirmed both of the above items.</p> <p>Describe your method for performing the 30-second timer accuracy check.</p> <p>The source is programmed for a 30-second exposure. This is timed with a stop watch. Criterion for acceptable result is agreement within one second.</p>			
Action Required/Taken: Document information above, then issue renewal			
Prepared By: SGabriel		Date: 3/28/05	