

March 29, 2005

Licensing Assistant Section  
Nuclear Materials Safety Branch  
U.S. Nuclear Regulatory Commission, Region 1  
475 Allendale Road  
King of Prussia, PA 19406-1415

J-1

06-02388-01

03001250

Dear Ms. Gabriel:

Re: Close-Out Survey of Old Hot Lab

On March 29, 2005, all radioactive material was relocated from the Hot Lab located in the old Radiation Oncology department to the new department located at 5 Highland St, New Britain, CT. As a result a close out survey has been performed. I have surveyed the room with an appropriate survey meter and have performed wipe tests in the room. The results showed no presence of residual radioactivity or dose rates above background in the room (see attachment for complete details of the close-out survey). This information is being forwarded to you so that the room may be released for unrestricted use.

Please contact me at your earliest convenience to update me on the status of this issue.

If you have any questions, please don't hesitate to contact me.

Sincerely,

George Pavlonnis, III M.S. DABR  
American Board of Radiology Certified Medical Physicist

(Ref: 136346)

136689  
NMSS/RGNI MATERIALS-002

### CLOSE-OUT SURVEY


Location Hot Lab - Old Radiation Oncology Department  
 Reason for Survey Close-Out Survey  
 Date of Survey March 29, 2005  
  
 Survey Meter Victoreen 290 (S/N 1116)  
 Date of Calibration July 20, 2004  
 Source Check 2.1 mr/hr (Acceptable Range: 1.6 to 2.4 mr/hr)  
 Background 0.02 mr/hr

Survey Performed by: *George Pavlounis* 3/29/05  
 George Pavlounis M.S. DABR

**Survey Results**

The entire room was surveyed with the equipment described above. All measurements at background levels.

Hot Lab Decommissioning Reference Map



Note: Not Drawn to Scale

**Wipe Test Results** (Captus 3000 S/N CNV-325)

Location	Total Counts	Efficiency	Activity (DPM)
Wipe #1	474	65%	Bkg
Wipe #2	474	65%	Bkg
Wipe #3	516	65%	Bkg
Wipe #4	485	65%	Bkg
Wipe #5	485	65%	Bkg
Wipe #6	488	65%	Bkg
Wipe #7	449	65%	Bkg
Wipe #8	490	65%	Bkg
Wipe #9	493	65%	Bkg
Wipe #10	513	65%	Bkg
Background	522	-	-

This is to acknowledge the receipt of your letter/application dated

3/29/2005, and to inform you that the initial processing which includes an administrative review has been performed.

Amendment 06-02388-C1  
There were no administrative omissions. Your application was assigned to a technical reviewer. Please note that the technical review may identify additional omissions or require additional information.

Please provide to this office within 30 days of your receipt of this card

---

A copy of your action has been forwarded to our License Fee & Accounts Receivable Branch, who will contact you separately if there is a fee issue involved.

Your action has been assigned **Mail Control Number** 136689.  
When calling to inquire about this action, please refer to this control number.  
You may call us on (610) 337-5398, or 337-5260.

BETWEEN: : (FOR LFMS USE)  
 : INFORMATION FROM LTS  
 : -----  
 :  
 License Fee Management Branch, ARM : Program Code: 02230  
 and : Status Code: 0  
 Regional Licensing Sections : Fee Category: 7C 2B 3E  
 : Exp. Date: 20120430  
 : Fee Comments: CODE 23  
 : Decom Fin Assur Reqd: N  
 : .....

LICENSE FEE TRANSMITTAL

A. REGION I

1. APPLICATION ATTACHED  
 Applicant/Licensee: NEW BRITAIN GENERAL HOSPITAL  
 Received Date: 20050329  
 Docket No: 3001250  
 Control No.: 136689  
 License No.: 06-02388-01  
 Action Type: Amendment

2. FEE ATTACHED  
 Amount:             
 Check No.:           

3. COMMENTS  
 Ref. 136 346  
 Signed Rebecca J. Jones  
 Date 3/31/2005

B. LICENSE FEE MANAGEMENT BRANCH (Check when milestone 03 is entered /\_\_/)

1. Fee Category and Amount: \_\_\_\_\_  
 2. Correct Fee Paid. Application may be processed for:  
 Amendment \_\_\_\_\_  
 Renewal \_\_\_\_\_  
 License \_\_\_\_\_  
 3. OTHER \_\_\_\_\_  
 \_\_\_\_\_

Signed \_\_\_\_\_  
 Date \_\_\_\_\_