

**From:** Sandra Gabriel  
**To:** WButler@WheelingHospital.com  
**Date:** Fri, Apr 1, 2005 3:51 PM  
**Subject:** NRC license amendment request for Wheeling Hospital, mail control 136537

Reference: Wheeling Hospital, Inc., 47-05322-02, 03012570, mail control 136537

Wayne Butler, Ph.D.:

One additional item:

I just checked the Sealed Source and Device Registry sheet for the Nucletron Model 105.999 afterloader. Nucletron Model 105.002 is the only approved source; Nucletron Model 096.001 and the Alpha-Omega source are approved only for your current Nucletron microSelectron-HDR Classic afterloader.

Item 7.E. of your license will be written as:

"Sealed Sources (Nucletron Model 105.002 [manufactured by Mallinckrodt Medical BV or AEA Technology, Inc.])"

Please contact me with any questions.

Sandy Gabriel

**Mail Envelope Properties**

(424DB453.11B : 15 : 11102)

**Subject:** NRC license amendment request for Wheeling Hospital, mail control  
136537  
**Creation Date:** Fri, Apr 1, 2005 3:51 PM  
**From:** Sandra Gabriel  
**Created By:** SLG2@nrc.gov

**Recipients**

WheelingHospital.com  
WButler (WButler@WheelingHospital.com)

**Post Office**

**Route**

WheelingHospital.com

**Files**

MESSAGE

**Size**

1442

**Date & Time**

Friday, April 1, 2005 3:51 PM

**Options**

**Expiration Date:** None  
**Priority:** Standard  
**Reply Requested:** No  
**Return Notification:** None

**Concealed Subject:** No  
**Security:** Standard