

ST. FRANCIS  
*Medical Center*



601 HAMILTON AVENUE • TRENTON, NJ 08629-1986

RECEIVED  
REGION I

03 MAR 24 12:44

March 18, 2005

United States Nuclear Regulatory Commission  
Region I  
475 Allendale Road  
King of Prussia, PA 19406-1415

K-8

Attention: Medical Licensing Division

Re: Materials License 29-00232-08


03002423

Dear Sir or Madam:

St. Francis Medical Center, Radioactive Materials License 29-00232-08, has relocated our Nuclear Medicine Department. The area where the department had been located was wipe tested, surveyed, and released for unrestricted use. A copy of the final survey is attached. Also enclosed is a copy of the floor plan of the new department.

If you have any questions or require additional information, please do not hesitate to contact myself or our consultant Health Physicist, Karen Wheeler, at (908) 788-9440, ext. 45.

Sincerely,

  
Joseph Flamini  
Vice President



A Member of Catholic Health East, Sponsored by Hope Ministries

136 43 3

NMCS/RGNI MATERIALS-002


SAINT FRANCIS MEDICAL CENTER  
601 Hamilton Avenue, Trenton, NJ 08629  
DEPARTMENT OF NUCLEAR MEDICINE

R/F Room 4

*Above: Security Dept. and ER*


*Below: Slab*

Radiology Work Core



SAINT FRANCIS MEDICAL CENTER  
601 Hamilton Avenue, Trenton, NJ 08629  
DEPARTMENT OF NUCLEAR MEDICINE  
Close-Out Survey

Corridor



**St. Francis Medical Center**  
**Department of Nuclear Medicine**  
 Close-out Survey and wipe tests

Instrumentation: Captus 3000 Well counter  
 Ludlum Model 3 Geiger counter with 44-9 probe  
 S/N 79264  
 Date of Calibration: 12/22/2004

Date of Survey: March 4, 2005

Location	Map Key	Survey (mR/hr)	Wipe Test (net cpm)
<i>Hot Lab:</i>			
Doorway, floor	1	0.03	9
Floor	2 - 4	0.03	44
Counters	5 - 7	0.04	9
Interior Door	8	0.03	24
Flood Source Storage	9-10	0.04	24
Base Cabinet, Counter (front)	11	0.03	34
Base Cabinet, Counter (center)	12	0.04	34
Base Cabinet, Counter (back)	13	0.03	5
Base Cabinet, Counter (back)	14	0.04	5
Base Cabinet, Counter (front)	15	0.03	19
<i>General Department:</i>			
Corridor	A - C	0.03	25
Office	D - F	0.04	27
Camera Room 1	G - I	0.04	36
Camera Room 2	J - L	0.03	22
Corridor, Ultrasound	M - O	0.03	32
<b>Background:</b>		0.03	261