

U.S. NUCLEAR REGULATORY COMMISSION		Conversation Date: February 10, 2005	
TELEPHONE CONVERSATION RECORD		Time: 11:15 am	
Mail Control No.: 022429	License No.: 42-21454-01E	Docket No.: 030-20785	
Licensee/Applicant Participant(s): Rhonda Harris, RSO	Organization: Ludlum Measurements, Inc.	Telephone No.: 1-800-622-0828	
Person(s) Calling: MSIB: Jonathan Rivera, Anthony Kirkwood			
Subject: Clarifications to deficiency response dated January 25, 2005, related to renewal of license			
<p>Summary:</p> <p>The licensee provided the following clarifications:</p> <ol style="list-style-type: none"> Counting standards and check sources to be distributed are sealed sources in solid form. Besides the isotopes in 10 CFR 30.71, Schedule B, there is the possibility that the licensee will continue to distribute radiation measuring instruments containing Am-241 to persons exempt from licensing pursuant to 10 CFR 30.15. Therefore the licensee desires to continue to be licensed to distribute these radiation measuring instruments. Counting standards and check sources are manufactured by outside companies, mainly Spectrum Techniques and Eberline. Licensee will provide NRC information on authorization of all source manufacturers to distribute the sources. All (100%) of radiation measuring instruments to be distributed are visually inspected and made sure screws on source holder are tight. The label, as provided in Attachments 8 and 9 of deficiency response dated January 25, 2005, would be placed on the source holder cover, as illustrated by Item 11 and Item 2 on Source Holder Assembly drawing provided in Attachment 4 of deficiency response. 			
Action Required/Taken:			
Prepared By: Jonathan Rivera		Date: February 10, 2005	