

U.S. NUCLEAR REGULATORY COMMISSION		Conversation Date: 4/11/2005	
TELEPHONE CONVERSATION RECORD		Time: 9:45 am	
Mail Control No.:	License No.:	Docket No.:	
136225	29-17610-01	030-13049	
Licensee/Applicant Participant(s):	Organization:	Telephone No.:	
Paula Visconti, Ph.D. (RSO)	Virtua Mem. Hosp. of Burlington Co.	609-261-7074	
Person(s) Calling: Joe Nick			
Subject: Request for Additional Information			
<p>Summary:</p> <p>I asked the licensee's representative for additional information regarding the recent application for a material license (dated 12/27/04).</p> <ol style="list-style-type: none"> 1. The application did not request authorization for <i>in vitro</i> kits per 31.11. Dr. Visconti stated that they no longer use these kits and the authorization is not required. 2. The application did not request authorization for the use of Cs-137 sources, authorized per 35.400, for the calibration of instruments. I asked if the licensee still wanted to retain this authorization? Dr. Visconti stated that they do not calibrate their instruments in-house (they are sent to a vendor for calibration), but they would like to keep this as an available option. Dr. Visconti also stated that she would be the authorized non-medical user for these sources.. 			
Action Required/Taken: Prepare the license renewal for issuance.			
Prepared By: Joe Nick		Date: 4/11/2005	