

TECEIVED DEGION 1

*05 MAR 14 P12:46

U.S. Nuclear Regulatory Commission Region 1 475 Allendale Road King of Prussia, PA 19406-1415 License # 07-28657-01 3/2/2005 03032378

Dear Sir or Madam,

I am submitting a proposal of changes as a result of a recent addition of an MRI Suite at our facility. Listed below are those modifications:

- 1. Moved the lead cabinetry in the Hot Lab
- 2. Reduced the size of the Hot Lab in order to build a hallway to the MRI Suite (see diagram)
- 3. Moved the Thyroid Probe / Well Counter into the new section (see diagram)
- 4. Moved the treadmill and EKG monitor into the new section (see diagram)

Understanding the importance of Radiation Safety, I have listed below the precautions that were taken before and during construction:

- 1. Thorough area surveys and wipes were done on 11/23/04, one day prior to moving the lead cabinetry, demolition of the pre-existing wall and reconstruction of the new wall
- 2. The Hot Lab remained secured at all times. Construction crews were only allowed access to the Hot Lab when the Nuclear Medicine Technologist was present. The workers were at no time exposed to any isotopes
- 3. The Co-57 sheet source remained in the lead case which the Nuclear Medicine Technologist removed from the Hot Lab while construction crews were actually in the Hot Lab
- 4. All sealed sources and e-vials (Co-57 and Cs-137) remained locked in the lead cabinet in the Hot Lab

710 South Queen Street
Dover, Delaware 19904
302-734-9888


136733 NMSS/RGNI MATERIALS-002 At this time I would also like to advise you that since our last NRC Inspection on 5/18/2004, we have purchased the Atomlab 930 Thyroid Probe and Well Counter which has replaced the Thyroid Spectrometer and the Victoreen Deluxe Wipe Test Counter.

Enclosed you will find copies of both the old and new area surveys. As before mentioned, I am submitting these changes as a proposal for your review and would appreciate any suggestions that you may have.


Respectfully Submitted,


Mahendra Parikh, MD

RSO, Mid-Delaware Imaging



MID-DELAWARE IMAGING NUCLEAR MEDICINE - AREA SURVEY FLOOR PLAN





MID-DELAWARE IMAGING NUCLEAR MEDICINE DAILY AREA SURVEYS & WEEKLY WIPES

AREA	SURVEY mR/hr	WIPES dpm	S DECONTAMIN	NATION	
AREA (1) RESTRICTED					
(A) HOT LAB COUNTER	}	(1)			
(B) L-SHIELD		(2)			
(C) HOT SHARPS	-	(3)			
(D) H.L. FLOOR		(4)			
(E) H.L. TRASH		(5)			
(F) H.L. DOOR KNOB		(6)		······································	
AREA (2) UNRESTRIC	TED				
(G) HAND SINK		(1)	April Marie Company		
(H) GAMMA CAMERA		(2)			
(I) SPECT TABLE		(3)			
(J) W.B. TABLE		(4)			
(K) TOSHIBA CONSOL	.E	(5)			
(L) INJECTION STAND		(6)			
(M) FLOOR-INJ. AREA		(7)			
(N) ROOM TRASH		(8)			
(O) TREADMILL	•	(9)			
(P) EKG CONSOLE		(10)		-	
SURVEY INSTRUMENT:			BKG:	mR/ hr	
WIPES INSTRUMENT:			BKG:	DPM	
DATE:	TECH	NOLOGIST:			
ACTION LEVELS					
	SURVEYS 5.0 mR/hr		<u>WIPES</u> 20,000 dpm (Tc99m) 2,000 dpm (I 123)		
UNRESTRICTED).5 mR/hr		2,000 dpm (Tc99m) 200 dpm (I 123)		

MID-DELAWARE IMAGING NUCLEAR THE DICINE DAILY AREA SURVEY

WIPE TEST

SURVEY

<2kdpm/100cm²

AREA	(mH/nr)	PASS	FAIL	DECON	NI AMINATION
STORAGE MODULE					·
HOT LAB COUNTER					
REFRIGERATOR			٠.		
L-SHIELD					
HOT SHARPS					
COLD SHARPS					
HOT LAB FLOOR					
HOT LAB TRASH HOT LAB DOOR					
HAND SINK				:	
GAMMA CAMERA					
SPECT TABLE					
WHOLE BODY TABLE					
EKG CONSOLE					
TREADMILL					
INJECTION STAND					
FLOOR - INJ AREA					
TRASH - INJ AREA					
DOOR-NUC. MED. DEPT.	•				
TOSHIBA CONSOLE		<u></u>			
TECH	•		ACTIO	N LEVELS	
WOTO INICTOLINATINT.		SWI	PES: 1-15	23 IN-111	Tc99m TI-201
VIPES INSTRUMENT:			1-131		C0-57/Ga- 67
CKGROUND	dpm	UNREST	RICTED: 2	00dpm	2,000dpm
				Odom	20,000 dp m
		RESTRIC	1ED: 2,00	0dpm	20,000 ap
RVEYINSTRUMENT			SURVEYS:		
AVEY NO MENT					
CKGROUND:	_mR/hr	UNRES	STRICTED:	.5mR/hr	
		RES	TRICTED:	5.0mR/hr	
*		i			

This is to acknowledge the	receipt of your letter/application dated
includes an administrative	, and to inform you that the initial processing which review has been performed.
	(アー) 身 しら アーレー rative omissions. Your application was assigned to a asse note that the technical review may identify additional ditional information.
Please provide to this o	ffice within 30 days of your receipt of this card
	peen forwarded to our License Fee & Accounts Receivable ou separately if there is a fee issue involved.
	out this action, please refer to this control number. 337-5398, or 337-5260.
NRC FORM 532 (RI) (6-96)	Sincerely, Licensing Assistance Team Leader

	: (FOR LFMS USE) : INFORMATION FROM LTS
BETWEEN:	:
License Fee Management Branch, ARM and Regional Licensing Sections	: Program Code: 02200 : Status Code: 0 : Fee Category: 7C : Exp. Date: 20120131 : Fee Comments: : Decom Fin Assur Reqd: N
LICENSE FEE TRANSMITTAL	
A. REGION I	
1. APPLICATION ATTACHED Applicant/Licensee: MID-DELAWARE II Received Date: 20050314 Docket No: 3032378 Control No.: 136732 License No.: 07-28657-01 Action Type: Amendment	MAGING, INC.
2. FEE ATTACHED Amount: Check No.:	
3. COMMENTS Signed _ Date _	Reletta funda
B. LICENSE FEE MANAGEMENT BRANCH (Check	when milestone 03 is entered //)
1. Fee Category and Amount:	
2. Correct Fee Paid. Application may Amendment Renewal License	be processed for:
3. OTHER	
Signed Date	