



UNITED STATES
NUCLEAR REGULATORY COMMISSION
REGION I
475 ALLENDALE ROAD
KING OF PRUSSIA, PENNSYLVANIA 19406-1415

April 26, 2005

Docket No. 03003094

License No. 37-07722-04

Brenda DeFeo
Vice President
The Bryn Mawr Hospital
130 South Bryn Mawr Avenue
Bryn Mawr, PA 19010

SUBJECT: INSPECTION 03003094/2005001, THE BRYN MAWR HOSPITAL, BRYN MAWR, PENNSYLVANIA SITE AND NOTICE OF VIOLATION

Dear Ms. DeFeo:

On April 13, 2005, Michelle Beardsley of this office conducted a safety inspection at the above address of activities authorized by the above listed NRC license. The inspection was an examination of your licensed activities as they relate to radiation safety and to compliance with the Commission's regulations and the license conditions. The inspection consisted of observations by the inspector, interviews with personnel, and a selected examination of representative records. The findings of the inspection were discussed with you, and members of your organization at the conclusion of the inspection.

Based on the results of this inspection, it appears that your activities were not conducted in full compliance with NRC requirements. A Notice of Violation is enclosed that categorizes the violation by severity level in accordance with NRC Enforcement Policy. You are required to respond to this letter and should follow the instructions specified in the enclosed Notice when preparing your response. In your response, you should document the specific actions taken and any additional actions you plan to prevent recurrence. Your response may reference or include previous docketed correspondence, if the correspondence adequately addresses the required response. After reviewing your response to this Notice, including your proposed corrective actions and the results of future inspections, the NRC will determine whether further NRC enforcement action is necessary to ensure compliance with NRC regulatory requirements.

Current NRC regulations and guidance are available at the NRC Web sites listed below or by contacting the Government Printing Office (GPO) toll-free at 1-888-293-6498. The GPO is open from 7:00 a.m. to 9:00 p.m. EST, Monday through Friday (except Federal holidays).

B. DeFeo
The Bryn Mawr Hospital

2

Your cooperation with us is appreciated.

Sincerely,

Original signed by Pamela J. Henderson

Pamela J. Henderson, Chief
Medical Branch
Division of Nuclear Materials Safety

Enclosure:
Notice of Violation

NRC Web site addresses
NRC regulations

<http://www.nrc.gov/reading-rm/doc-collections/cfr/>

Licensing guidance

<http://www.nrc.gov/reading-rm/doc-collections/nuregs/staff/sr1556/>

General Policy and Procedure for NRC Enforcement Actions

<Http://www.nrc.gov/what-we-do/regulatory/enforcement/enforc-pol.pdf>

206 of the Energy Reorganization Act of 1974

<http://www.nrc.gov/who-we-are/governing-laws.html>

cc:
Marchello Barbarisi, M.D., Radiation Safety Officer
Commonwealth of Pennsylvania

B. DeFeo
The Bryn Mawr Hospital

3

Distribution:
D. J. Holody, RI

DOCUMENT NAME: E:\Filenet\ML051180238.wpd

To receive a copy of this document, indicate in the box: "C" = Copy w/o attach/encl "E" = Copy w/ attach/encl "N" = No copy

OFFICE	DNMS/RI	N	DNMS/RI		DNMS/RI			
NAME	MBeardsley		PHenderson					
DATE	4/26/05		4/26/05					

OFFICIAL RECORD COPY

NOTICE OF VIOLATION

The Bryn Mawr Hospital
Bryn Mawr, PA

Docket No. 03003094
License No. 37-07722-04

During an NRC inspection conducted on April 13, 2005, one violation of NRC requirements was identified. In accordance with the NRC Enforcement Policy, the violation is listed below:

Condition No. 15 of License No. 37-07722-04 requires, in part, that the licensee shall conduct its program in accordance with statements, representations, and procedures contained in the application dated September 26, 1994.

Item 10.4 of the application dated September 26, 2004, states that the licensee has implemented the model safety rules published in Appendix I to Regulatory Guide 10.8, Rev. 2. Item 6 of Appendix I to Regulatory Guide 10.8, Rev. 2 states that the licensee will not store food, drinks or personal effects in areas where radioactive material is stored or used.

Contrary to the above, on April 13, 2005, the licensee stored drinks in an area where radioactive material is stored. Specifically, a drink was found by the inspector in the area where radioactive material is stored in a refrigerator.

This is a Severity Level violation IV (Supplement VI).

Pursuant to the provisions of 10 CFR 2.201, The Bryn Mawr Hospital is hereby required to submit a written statement or explanation to the U.S. Nuclear Regulatory Commission, ATTN: Document Control Desk, Washington, D.C. 20555, with a copy to the Regional Administrator, Region I, within 30 days of the date of the letter transmitting this Notice of Violation (Notice). This reply should be clearly marked as a "Reply to a Notice of Violation" and should include for each violation: (1) the reason for the violation, or, if contested, the basis for disputing the violation, (2) the corrective steps that have been taken and the results achieved, (3) the corrective steps that will be taken to avoid further violations, and (4) the date when full compliance will be achieved. Your response may reference or include previous docketed correspondence, if the correspondence adequately addresses the required response. If an adequate reply is not received within the time specified in this Notice, an order or a Demand for Information may be issued as to why the license should not be modified, suspended, or revoked, or why such other action as may be proper should not be taken. Where good cause is shown, consideration will be given to extending the response time.

If you contest this enforcement action, you should also provide a copy of your response to the Director, Office of Enforcement, United States Nuclear Regulatory Commission, Washington, DC 20555-0001. Under the authority of Section 182 of the Act, 42 U.S.C. 2232, any response which contests an enforcement action shall be submitted under oath or affirmation.

Your response will be placed in the NRC Public Document Room (PDR) and on the NRC Web site at <http://www.nrc.gov/reading-rm.html>. To the extent possible, it should, therefore, not include any personal privacy, proprietary, or safeguards information so that it can be made publically available without redaction. However, if you find it necessary to include such

Notice of Violation
The Bryn Mawr Hospital

2

information, you should clearly indicate the specific information that you desire not to be placed in the PDR, and provide the legal basis to support your request for withholding the information from the public.

In accordance with 10 CFR 19.11, you may be required to post this Notice within two working days.

Dated This 26 day of April 2005