

**SOUTHWEST RESEARCH INSTITUTE**

**Department of Quality Assurance  
Calibration Laboratory**

**CERTIFICATE OF CALIBRATION  
11/08/94**

Issued to: NARASI SRIDHAR DIV20 ,B57  
Manufacturer: SWRI  
Nomenclature: RESISTOR BOX  
Serial Number: 171002

Asset Number: 003095  
Model Number: 10 OHM  
SwRI Capital Number: NONE

**ENVIRONMENTAL CONDITIONS**

Temperature: 70.0F

Relative Humidity: 38 %

**CALIBRATION INFORMATION**

Location: CAL1  
Procedure Number: SWRI  
Remarks: ACCURACY IS +/- 0.1 OHMS(+/- 1%)  
MEASURED VALUE IS 10.06 OHMS

Technician: 10004  
Accuracy: +/-1%  
Received IN Tolerance

**Calibration was in accordance with requirements of MIL-STD-45662A.  
Measurements are traceable to the National Institute of Standards and Technology.  
Inspection and test data are on file and available for inspection.**

**STANDARDS USED FOR CERTIFICATION**

Asset #	Serial #	Mfg	Model #	Nomenclature	Cal Date	Int.	Cal Due
001505	2823A07741	HP	3458A	SYSTEM MULTIMETER	05/04/94	12	05/04/95

Certified by : Walt Stiv

Certificate#: 15596

Calibration Date: 11/08/94  
Interval: 6 months  
Next Calibration Due: 05/08/95

**SOUTHWEST RESEARCH INSTITUTE**  
**Department of Quality Assurance**  
**Calibration Laboratory**

**CERTIFICATE OF CALIBRATION**  
**05/11/95**

Issued to: DARRELL DUNN  
Manufacturer: SWRI  
Nomenclature: RESISTOR BOX  
Serial Number: 171002

DIV20 ,B57

Asset Number: 003095  
Model Number: 10 OHM

SwRI/Div. I.D. #: NONE

**ENVIRONMENTAL CONDITIONS**

Temperature: 70.0F

Relative Humidity: 48%

**CALIBRATION INFORMATION**

Procedure Number: SWRI

Accuracy: +/-1%

Remarks: MEASURED RESISTANCE IS 10.068 OHMS

Received IN Tolerance

**Calibration was in accordance with requirements of MIL-STD-45662A. Measurements are traceable to the National Institute of Standards and Technology. Inspection and test data are on file and available for inspection.**

**STANDARDS USED FOR CERTIFICATION**

Asset #	Serial #	Mfg	Model #	Nomenclature	Cal Date	Int. Cal Due
000115	2201A08306	HP	3456A	DIGITAL VOLTMETER	11/29/94 6	05/29/95

Certified by :



Certificate#: 17392

Calibration Date: 05/11/95  
Interval: 6 months  
Next Calibration Due: 11/11/95

**SOUTHWEST RESEARCH INSTITUTE**  
**Department of Quality Assurance**  
**Calibration Laboratory**  
**CERTIFICATE OF CALIBRATION**  
**11/22/95**

Issued to: DARRELL DUNN DIV20 ,B57  
Manufacturer/Model: SWRI /10 OHM  
Nomenclature: RESISTOR BOX  
Serial Number: 171002  
Asset Number: 003095  
Notes:

**ENVIRONMENTAL CONDITIONS**

Temperature: 72.0F Relative Humidity: 42%

**CALIBRATION INFORMATION**

Procedure Number: SWRI Accuracy: +/-1%  
Remarks: Received IN Tolerance

**Calibration was in accordance with requirements of MIL-STD-45662A.**  
**Measurements are traceable to the National Institute of Standards and Tech-**  
**nology. Inspection and test data are on file and available for inspection.**

**STANDARDS USED FOR CERTIFICATION**

Asset #	Serial #	Mfg	Model #	Nomenclature	Cal Date	Int.	Cal Due
001505	2823A07741	HP	3458A	SYSTEM MULTIMETER	05/11/95	12	05/11/96

Certified by : *Anthony Plummer*

Certificate#: 19322

Calibration Date: 11/22/95  
Interval: 6 months  
Next Calibration Due: 05/22/96



Southwest Research Institute  
6220 Culebra Road  
San Antonio, TX 78238  
Department of Quality Assurance  
Calibration Laboratory

# Certificate of Calibration

14 June 1996

Issued to: DARRELL DUNN                      DIV20                      B57  
Manufacturer/Model: SWRI 10 OHM  
Description: RESISTOR BOX  
Serial Number: 171002  
Asset Number: 003095

## Environmental Conditions

Temperature: 73.0                                              Humidity: 44%

## Calibration Information

Calibration was in accordance with requirements of MIL-STD-45662A and ANSI/NC SL Z540-1-1994. Measurements are traceable to the National Institute of Standards and Technology (NIST). This report may not be reproduced except in full without written approval of the originator. Inspection and test data are on file and available for inspection.

Calibration Date: 13 Jun 96                                              Calibration Procedure: CLCP-OR-001  
Interval: 6 months                                              Accuracy: +/-1%  
Next Calibration Due: 13 Dec 96                                              Received: In Tolerance

Remarks:

Certificate # 21450

Signed: 

LAST PAGE OF REPORT  
Total Pages Printed: 1



Southwest Research Institute  
6220 Culebra Road  
San Antonio, TX 78238  
Department of Quality Assurance  
Calibration Laboratory

# Certificate of Calibration

17 December 1996

Issued to: DARRELL DUNN                      DIV20                      B57  
Manufacturer/Model: SWRI 10 OHM  
Description: RESISTOR BOX  
Serial Number: 171002  
Asset Number: 003095

## Environmental Conditions

Temperature: 70.0 Deg. F                      Humidity: 34%

## Calibration Information

Calibration was in accordance with requirements of MIL-STD-45662A and ANSI/NCSL Z540-1-1994. Measurements are traceable to the National Institute of Standards and Technology (NIST). This report may not be reproduced except in full without written approval of the originator. Inspection and test data are on file and available for inspection.

Calibration Date: 17 Dec 96                      Calibration Procedure: CLCP-OR-001  
Interval: 6 months                      Accuracy: +/-1%  
Next Calibration Due: 17 Jun 97                      Received: In Tolerance

Remarks: AVERAGE MEASURED VALUE OVER 10 SAMPLES:  
10.06745 OHMS

## Standards Used

Asset	MFR	Model	Description	Serial No.	Due Cal
001505	HP	3458A	SYSTEM MULTIMETER	2823A07741	18 Jan 97

Certificate # 23527

Signed: 

LAST PAGE OF REPORT  
Total Pages Printed: 1



Southwest Research Institute  
6220 Culebra Road  
San Antonio, TX 78238  
Department of Quality Assurance  
Calibration Laboratory

# Certificate of Calibration

21 July 1997

Issued to: DARRELL DUNN                      DIV20                      B57  
Manufacturer/Model: SWRI 10 OHM  
Description: RESISTOR BOX  
Serial Number: 171002  
Asset Number: 003095

## Environmental Conditions

Temperature: 74.0 Deg. F                      Humidity: 47%

## Calibration Information

Calibration was in accordance with requirements of MIL-STD-45662A and ANSI/NCSL Z540-1-1994. Measurements are traceable to the National Institute of Standards and Technology (NIST). This report may not be reproduced except in full without written approval of the originator. Inspection and test data are on file and available for inspection.

Calibration Date: 21 Jul 97                      Calibration Procedure: CLCP-OR-001  
Interval: 6 months                      Accuracy: +/-1%  
Next Calibration Due: 21 Jan 98                      Received: In Tolerance

Remarks: AVERAGE OHM VALUE OVER 10 SAMPLES:  
10.06805 OHMS

## Standards Used

Asset	MFR	Model	Description	Serial No.	Due Cal
001505	HEWLETT-PA	3458A	SYSTEM MULTIMETER	2823A07741	1 Feb 98

Certificate # 26041

Signed:

LAST PAGE OF REPORT  
Total Pages Printed: 1





Southwest Research Institute  
6220 Culebra Road  
San Antonio, TX 78238  
(210) 522-5215  
Department of Quality Assurance  
Calibration Laboratory

Accredited



Certificate #  
0972-01

## Certificate of Calibration

1 April 1999

**Issued to:** DARRELL DUNN DIV20 B57  
**Manufacturer/Model:** SWRI 10 OHM  
**Description:** RESISTOR BOX  
**Serial Number:** 171002  
**Asset Number:** 003095

This certifies the above item was calibrated in compliance with MIL-STD-45662A and ANSI/NCSL Z540-1-1994. Standards used in this calibration, described in the referenced calibration procedure with associated uncertainties or tolerances, are traceable to the National Institute of Standards and Technology (NIST). Supporting documentation relative to traceability is on file and is available for examination upon request. This certificate is not to be reproduced, except in full, without the written approval of the Southwest Research Institute Department of Quality Assurance Calibration Laboratory.

This laboratory is accredited by the American Association for Laboratory Accreditation (A2LA) and the results of this calibration certificate were determined in accordance with the terms of accreditation unless stated otherwise below.

The uncertainty of the calibration was sufficient to determine that the item met the manufacturer's published specifications unless stated otherwise below.

**Ambient Conditions:** Temperature: 74.0 Degrees Fahrenheit Humidity: 40 % RH

**Calibration Date:** 1 Apr 99 **Calibration Procedure:** CLCP-OR-001

**Condition as Received:** IN TOLERANCE

**Condition as Released:** IN TOLERANCE

**Remarks:** MEASURED RESISTANCE VALUE: 10.08921 OHMS  
STANDARD MEASUREMENT UNCERTAINTY: .0061%

**Approved by:**

Jim Patterson, Supervisor or Walt Hill, Metrologist

**Certificate #** 33574

m:\a2la.rpt Rev date 10 Mar 99

**Measurements performed by:**

Vince Morales, Technician

Page 1 of 1





Southwest Research Institute  
6220 Culebra Road  
San Antonio, TX 78238  
(210) 522-5215  
Department of Quality Assurance  
Calibration Laboratory



Certificate #  
0972-01

## Certificate of Calibration

30 September 1999

**Issued to:** DARRELL DUNN DIV20 B57  
**Manufacturer/Model:** SWRI 10 OHM  
**Description:** RESISTOR BOX  
**Serial Number:** 171002  
**Asset Number:** 003095

This certifies the above item was calibrated in compliance with MIL-STD-45662A and ANSI/NCSL Z540-1-1994. Standards used in this calibration, described in the referenced calibration procedure with associated uncertainties or tolerances, are traceable to the National Institute of Standards and Technology (NIST). Supporting documentation relative to traceability is on file and is available for examination upon request. This certificate is not to be reproduced, except in full, without the written approval of the Southwest Research Institute Department of Quality Assurance Calibration Laboratory.

This laboratory is accredited by the American Association for Laboratory Accreditation (A2LA) and the results of this calibration certificate were determined in accordance with the terms of accreditation unless stated otherwise below.

The uncertainty of the calibration was sufficient to determine that the item met the manufacturer's published specifications unless stated otherwise below.

**Ambient Conditions:** Temperature: 74.0 Degrees Fahrenheit Humidity: 40 % RH


**Calibration Date:** 20 Sep 99      **Calibration Procedure:** CL-70, JUN 99

**Condition as Received:** IN TOLERANCE


**Condition as Released:** IN TOLERANCE

**Remarks:** AVERAGE MEASURED VALUE: 10.068 OHMS  
MEASUREMENT UNCERTAINTY: +/-1 MILLIOHM

**Approved by:**

  
\_\_\_\_\_  
Jim Patterson, Supervisor or Walt Hill, Metrologist

**Measurements performed by:**

  
\_\_\_\_\_  
Vince Morales, Technician

Certificate # 36050

Page 1 of 1



# SOUTHWEST RESEARCH INSTITUTE™

6220 Culebra Road, P.O. Drawer 28510  
Institute Quality Systems  
Institute Calibration Laboratory  
Phone: 210-522-5215 Fax 210-522-3692

## Certificate of Calibration

**Submitted By:** DIV20

**Address:** B57

**Contact:** DARRELL DUNN

**Manufacturer Model:** SWRI 10 OHM

**Description:** RESISTOR BOX

**Serial No:** 171002

**Asset No:** 003095

**Procedure:** CL-70, JUN 99

**Work Order:** 444050662

**Date Issued:** Oct 15, 2002

**Calibration Date:** Oct 15, 2002

**\*\*Calibration Due:** Apr 15, 2003

**Calibration Location:** N/A

**Environment:** Temp. 72.0°F Hum. 40 %RH

**\*As Found:** IN TOLERANCE

**\*As Left:** IN TOLERANCE

This certificate documents traceability to the National Institute of Standards and Technology (NIST) and the International System of Units (SI). The Laboratory quality system conforms to ISO/IEC 17025, 1999 and ANSI/NC SL Z540-1-1994 which are equivalent to relevant requirements of the ISO 9000-1994 series of standards. This certificate may not be reproduced, except in full, without the written approval of the Southwest Research Institute Calibration Laboratory. The results of this calibration relate only to the individual instrument described above. This certificate shall not be used to claim product endorsement by the American Association for Laboratory Accreditation (A2LA) or any agency of the U. S. Government.

Uncertainty evaluation includes the item under test and is calculated in accordance with the ISO "Guide to the Expression of Uncertainty in Measurement" (GUM). The uncertainty represents an expanded uncertainty using a coverage factor of  $k=2$  to approximate a 95% confidence level. The calibration process provides a Test Uncertainty Ratio (TUR) of less than or equal to 25% (4:1) of the test limit unless otherwise stated in remarks or an attachment.

\*The client has sole responsibility for determination of in/out of tolerance or compliance/noncompliance. An in/out of tolerance opinion is provided for your convenience based only on the Test Instrument (TI) reading(s) and limits as reported. The reported uncertainty relates only to the results at the time of calibration and does not imply any short or long term stability of the TI.

\*\*Calibration interval is determined by the client and does not assure the instrument will remain within tolerance until this date. Any number of factors may cause the instrument to be out of tolerance before the next calibration date.

**Remarks:** None

### Standards Used

Asset	Manufacturer	Model	Description	Cal Due
007001	HEWLETT-PACKARD	3458A/OPT-002	MULTIMETER	Jan 06, 03

Approved by: Walt Hill

Metrology Group Leader

m:\Non21a1.rpt Rev date 15, August 02

Measurements by: Tony Planas

Metrology Technician

Southwest Research Institute  
Calibration Laboratory  
Calibration Sheet

Work Order: 444050662	Mfr. SWRI	Technician Tplanas
Asset No. 3095	Model 10 OHM	Procedure CL-70 06/99
Serial No. 171002	Type. Resistor	Cal Date. 15-Oct-02
Remarks:		

Function/Range	Test Point	TI Reading	Difference	Test Limits+/-	Uncertainty	Found/Left
Resistance	Ohm	Ohm	Ohm	Ohm	Ohm	Results
	10.00	10.06	0.06	0.10	>4:1	Pass



# SOUTHWEST RESEARCH INSTITUTE™

6220 Culebra Road, P.O. Drawer 28510  
Institute Quality Systems  
Institute Calibration Laboratory  
Phone: 210-522-5215 Fax 210-522-3692

## Certificate of Calibration

**Submitted By:** DIV20

**Address:** B57

**Contact:** DARRELL DUNN

**Manufacturer Model:** SWRI 10 OHM

**Description:** RESISTOR BOX

**Serial No:** 171002

**Asset No:** 003095

**Procedure:** CL-70, JUN 99

**Work Order:** 444053024

**Date Issued:** Apr 3, 2003

**Calibration Date:** Apr 3, 2003

**\*\*Calibration Due:** Oct 3, 2003

**Calibration Location:** Bldg. 64

**Environment:** Temp. 68.0°F Hum. 42 %RH

**\*As Found:** IN TOLERANCE

**\*As Left:** IN TOLERANCE

This certificate documents traceability to the National Institute of Standards and Technology (NIST) and the International System of Units (SI). The Laboratory quality system conforms to ISO/IEC 17025, 1999 and ANSI/NCCL Z540-1-1994 which are equivalent to relevant requirements of the ISO 9000-1994 series of standards. This certificate may not be reproduced, except in full, without the written approval of the Southwest Research Institute Calibration Laboratory. The results of this calibration relate only to the individual instrument described above. This certificate shall not be used to claim product endorsement by the American Association for Laboratory Accreditation (A2LA) or any agency of the U. S. Government.

Uncertainty evaluation includes the item under test and is calculated in accordance with the ISO "Guide to the Expression of Uncertainty in Measurement" (GUM). The uncertainty represents an expanded uncertainty using a coverage factor of  $k=2$  to approximate a 95% confidence level. The calibration process provides a Test Uncertainty Ratio (TUR) of less than or equal to 25% (4:1) of the test limit unless otherwise stated in remarks or an attachment.

\*The client has sole responsibility for determination of in/out of tolerance or compliance/noncompliance. An in/out of tolerance opinion is provided for your convenience based only on the Test Instrument (TI) reading(s) and limits as reported. The reported uncertainty relates only to the results at the time of calibration and does not imply any short or long term stability of the TI.

\*\*Calibration interval is determined by the client and does not assure the instrument will remain within tolerance until this date. Any number of factors may cause the instrument to be out of tolerance before the next calibration date.

**Remarks:** None

### Standards Used

Asset	Manufacturer	Model	Description	Cal Due
007001	HEWLETT-PACKARD	3458A/OPT-002	MULTIMETER	Jan 17, 04

Approved by: Walt Hill

Metrology Group Leader

m:\Nona21a1.rpt Rev date 15, August 02

Measurements by: Tom Hannon

Metrology Technician

Southwest Research Institute  
 Calibration Laboratory  
 Calibration Sheet

Work Order: 444053024	Mfr. SWRI	Technician Thomas Hannon
Asset No. 003095	Model 10 OHM	
Serial No. 171002	Type. Resistor	Cal Date. 3-Apr-03
Remarks:		

Function/Range	Test Point	TI Reading	Difference	Test Limits+/-	Uncertainty	Found/Left
Resistance	Ohm 10.00	Ohm 10.07	Ohm 0.07	Ohm 0.10	Ohm 0.0008	Results Pass

END OF REPORT



# SOUTHWEST RESEARCH INSTITUTE™

6220 Culebra Road, P.O. Drawer 28510  
Institute Quality Systems  
Institute Calibration Laboratory  
Phone: 210-522-5215 Fax 210-522-3692



Certificate #

0972-01

## Certificate of Calibration

**Submitted By:** DIV20  
**Address:** B57  
**Contact:** DARRELL DUNN  
**Manufacturer Model:** SWRI 10 OHM  
**Description:** RESISTOR BOX  
**Serial No:** 171002  
**Asset No:** 003095  
**Procedure:** CL-70, JUN/99

**Work Order:** 444055696  
**Date Issued:** Oct 7, 2003  
**Calibration Date:** Oct 3, 2003  
**\*\*Calibration Due:** Apr 3, 2004  
**Calibration Location:** Bldg. 64  
**Environment:** Temp. 70.0°F Hum. 40 %RH  
**\*As Found:** IN TOLERANCE  
**\*As Left:** IN TOLERANCE

This certificate documents traceability to the National Institute of Standards and Technology (NIST) and the International System of Units (SI). The Laboratory quality system conforms to ISO/IEC 17025, 1999 and ANSI/NCSL Z540-1-1994 which are equivalent to relevant requirements of the ISO 9000-1994 series of standards. This certificate may not be reproduced, except in full, without the written approval of the Southwest Research Institute Calibration Laboratory. The results of this calibration relate only to the individual instrument described above. This certificate shall not be used to claim product endorsement by the American Association for Laboratory Accreditation (A2LA) or any agency of the U. S. Government.

Uncertainty evaluation includes the item under test and is calculated in accordance with the ISO "Guide to the Expression of Uncertainty in Measurement" (GUM). The uncertainty represents an expanded uncertainty using a coverage factor of  $k=2$  to approximate a 95% confidence level. The calibration process provides a Test Uncertainty Ratio (TUR) of less than or equal to 2.5% (4:1) of the test limit unless otherwise stated in remarks or an attachment.

\*The client has sole responsibility for determination of in/out of tolerance or compliance/noncompliance. An in/out of tolerance opinion is provided for your convenience based only on the Test Instrument (TI) reading(s) and limits as reported. The reported uncertainty relates only to the results at the time of calibration and does not imply any short or long term stability of the TI.

\*\*Calibration interval is determined by the client and does not assure the instrument will remain within tolerance until this date. Any number of factors may cause the instrument to be out of tolerance before the next calibration date.

**Remarks:** None

### Standards Used

Asset	Manufacturer	Model	Description	Cal Due
001505	HEWLETT-PACKARD	3458A/OPT-002	MULTIMETER	Apr 14, 04

Approved by: Walt Hill  
Metrology Group Leader

Measurements by: Vince Morales  
Metrology Technician

Southwest Research Institute  
 Calibration Laboratory  
 Calibration Report

Work Order:	444055696	Mfr.	SWRI	Technician	Vmorales
Asset No.	3095	Model	10 OHM		
Serial No.	171002	Type.	Resistor	Cal Date.	3-Oct-03
Remarks:					

Function/Range	Test Point	TI Reading	Difference	+/-Limit	+/-Uncertainty	Found/Left
Resistance	Ohm 10.00	Ohm 10.06	Ohm 0.06	Ohm 0.10	Ohm 0.0008	Results Pass

END OF REPORT



# SOUTHWEST RESEARCH INSTITUTE™

6220 Culebra Road, P.O. Drawer 28510  
Institute Quality Systems  
Institute Calibration Laboratory  
Phone: 210-522-5215 Fax 210-522-3692



Certificate #  
0972-01

## Certificate of Calibration

**Submitted By:** DIV20  
**Address:** B57  
**Contact:** DARRELL DUNN  
**Manufacturer Model:** SWRI 10 OHM  
**Description:** RESISTOR BOX  
**Serial No:** 171002  
**Asset No:** 003095  
**Procedure:** RESISTORS JAN/04

**Work Order:** 444058431  
**Date Issued:** Apr 7, 2004  
**Calibration Date:** Apr 7, 2004  
**\*\*Calibration Due:** Oct 7, 2004  
**Calibration Location:** Bldg. 64  
**Environment:** Temp. 73.0°F Hum. 45 %RH  
**\*As Found:** IN TOLERANCE  
**\*As Left:** IN TOLERANCE

This certificate documents traceability to the National Institute of Standards and Technology (NIST) and the International System of Units (SI). The Laboratory quality system conforms to ISO/IEC 17025, 1999 and ANSI/NCSL Z540-1-1994 which are equivalent to relevant requirements of the ISO 9000-1994 series of standards. This certificate may not be reproduced, except in full, without the written approval of the Southwest Research Institute Calibration Laboratory. The results of this calibration relate only to the individual instrument described above. This certificate shall not be used to claim product endorsement by the American Association for Laboratory Accreditation (A2LA) or any agency of the U. S. Government

Uncertainty evaluation includes the item under test and is calculated in accordance with the ISO "Guide to the Expression of Uncertainty in Measurement" (GUM). The uncertainty represents an expanded uncertainty using a coverage factor of k=2 to approximate a 95% confidence level. The calibration process provides a Test Uncertainty Ratio (TUR) of less than or equal to 25% (4:1) of the test limit unless otherwise stated in remarks or an attachment


\*The client has sole responsibility for determination of in/out of tolerance or compliance/noncompliance. An in/out of tolerance opinion is provided for your convenience based only on the Test Instrument (TI) reading(s) and limits as reported. The reported uncertainty relates only to the results at the time of calibration and does not imply any short or long term stability of the TI.

\*\*Calibration interval is determined by the client and does not assure the instrument will remain within tolerance until this date. Any number of factors may cause the instrument to be out of tolerance before the next calibration date.


**Remarks:** None

### Standards Used

Asset	Manufacturer	Model	Description	Cal Due
010465	FLUKE	8508A-01	MULTIMETER	Aug 27, 04

  
\_\_\_\_\_  
Approved by: Walt Hill

Metrology Group Leader  
m:\a2la1.rpt Rev date 15, August 02

  
\_\_\_\_\_  
Measurements by: Bob Trollinger

Metrology Technician



Southwest Research Institute  
Calibration Laboratory  
Calibration Report

Work Order:	444058431	Mfr.	SWRI	Technician	blt
Asset No.	003095	Model	10 OHM	Cal Date.	7-Apr-04
Serial No.	171002	Type.	Resistor		
Remarks:					

Function/Range	Test Point	TI Reading	Difference	+/-Limit	+/-Uncertainty	Found/Left
Resistance	Ohm	Ohm	Ohm	Ohm	Ohm	Results
G-W	10.00	10.07	0.07	0.10	0.0008	Pass
G-R	10.00	10.07	0.07	0.10	0.0008	Pass

END OF REPORT



# SOUTHWEST RESEARCH INSTITUTE™

6220 Culebra Road, P.O. Drawer 28510  
Institute Quality Systems  
Institute Calibration Laboratory  
Phone: 210-522-5215 Fax 210-522-3692



## Certificate of Calibration

Certificate #  
0972-01

**Submitted By:** DIV20  
**Address:** B57  
**Contact:** DARRELL DUNN  
**Manufacturer Model:** SWRI 10 OHM  
**Description:** RESISTOR BOX  
**Serial No:** 171002  
**Asset No:** 003095  
**Procedure:** RESISTORS JAN/04

**Work Order:** 444061101  
**Date Issued:** Oct 14, 2004  
**Calibration Date:** Oct 14, 2004  
**\*\*Calibration Due:** Apr 14, 2005  
**Calibration Location:** Bldg. 64  
**Environment:** Temp. 73.0°F Hum. 40 %RH  
**\*As Found:** IN TOLERANCE  
**\*As Left:** IN TOLERANCE

This certificate documents traceability to the National Institute of Standards and Technology (NIST) and the International System of Units (SI). The Laboratory quality system conforms to ISO/IEC 17025, 1999 and ANSI/NCCL Z540-1-1994 which are equivalent to relevant requirements of the ISO 9000-1994 series of standards. This certificate may not be reproduced, except in full, without the written approval of the Southwest Research Institute Calibration Laboratory. The results of this calibration relate only to the individual instrument described above. This certificate shall not be used to claim product endorsement by the American Association for Laboratory Accreditation (A2LA) or any agency of the U. S. Government.

Uncertainty evaluation includes the item under test and is calculated in accordance with the ISO "Guide to the Expression of Uncertainty in Measurement" (GUM). The uncertainty represents an expanded uncertainty using a coverage factor of k=2 to approximate a 95% confidence level. See Remarks or attached Calibration Report with the same Work Order number for calibration data.

\*The client has sole responsibility for determination of in/out of tolerance or compliance/noncompliance. An in/out of tolerance opinion is provided for your convenience based only on the Test Instrument (TI) reading(s) and limits as reported. The reported uncertainty relates only to the results at the time of calibration and does not imply any short or long term stability of the TI.

\*\*Calibration interval is determined by the client and does not assure the instrument will remain within tolerance until this date. Any number of factors may cause the instrument to be out of tolerance before the next calibration date.

**Remarks:** None

### Standards Used

Asset	Manufacturer	Model	Description	Cal Due
007001	HEWLETT-PACKARD	3458A/OPT-002	MULTIMETER	Feb 19, 05

Approved by: Walt Hill  
Metrology Group Leader  
m:\a2la1.rpt Rev date 11, May 04

Measurements by: Scott Kester  
Metrology Technician

Southwest Research Institute  
 Calibration Laboratory  
 Calibration Report

Work Order:	444061101	Mfr.	SWRI	Technician	SRK
Asset No.	003095	Model	10 OHM	Cal Date.	14-Oct-04
Serial No.	171002	Type.	Resistor		
Remarks:					

Function/Range	Test Point	TI Reading	Difference	+/-Limit	+/-Uncertainty	Found/Left
Resistance	Ohm	Ohm	Ohm	Ohm	Ohm	Results
G-W	10.00	10.08	0.08	0.10	0.00075	Pass
G-R	10.00	10.08	0.08	0.10	0.00075	Pass

END OF REPORT