

S O U T H W E S T R E S E A R C H I N S T I T U T E

Department of Quality Assurance
Calibration Laboratory

CERTIFICATE OF CALIBRATION

Issued to: DIV20 B57 NARASI SRIDHAR

Device No: 2472

Manufacturer: FISHER

Model: 14-983-10B

Nomenclature: THERMOMETER

Serial Number: 183306

SwRI No: NONE

Cal interval 6 Mo.

Remarks

Accuracy: +/- 1DEG.C

Procedure: SWRI

ENVIRONMENT

Temperature:

Humidity:

Location: SWRI DIV04 BLDG. 145

CONCLUSION

Tolerance/Remarks: Received in tolerance, no adjustments made

Room temperature.

Calibration was in accord with requirements of MIL-STD-45662A. Measurements are traceable to the National Institute of Standards and Technology. Inspection and test data are on file and available for inspection.

Signed

Calibration Date: 01/11/94

Cal interval: 6 Months

Record Number: 00013050

Next Calibration Due: 07/11/94

S O U T H W E S T R E S E A R C H I N S T I T U T E

Department of Quality Assurance
Calibration Laboratory

Device Serial No: 183306

Calibration Date: 01/11/94

STANDARDS

Standard No: 219 Manufacturer: AZONIX

Model: A1011

Nomenclature: RTD THERMOMETER

Serial No: T1587-2078

Cal.Due: 02/12/94

Cal.Rec.No: 00010541

Standard No: 328 Manufacturer: AZONIX

Model: A12001

Nomenclature: RTD TEMPERATURE PROBE

Serial No: 351477

Cal.Due: 02/12/94

Cal.Rec.No: 00010540