



Southwest Research Institute
6220 Culebra Road
San Antonio, TX 78238
(210) 522-5215
Department of Quality Assurance
Calibration Laboratory



Certificate #
0972-01

Certificate of Calibration

17 March 1999

Issued to: DARRELL DUNN DIV20 B57
Manufacturer/Model: FISHER SCIENTIFIC 15-166A
Description: THERMOMETER
Serial Number: C98-132
Asset Number: 007170

This certifies the above item was calibrated in compliance with MIL-STD-45662A and ANSI/NCSL Z540-1-1994. Standards used in this calibration, described in the referenced calibration procedure with associated uncertainties or tolerances, are traceable to the National Institute of Standards and Technology (NIST). Supporting documentation relative to traceability is on file and is available for examination upon request. This certificate is not to be reproduced, except in full, without the written approval of the Southwest Research Institute Department of Quality Assurance Calibration Laboratory.

This laboratory is accredited by the American Association for Laboratory Accreditation (A2LA) and the results of this calibration certificate were determined in accordance with the terms of accreditation unless stated otherwise below.

The uncertainty of the calibration was sufficient to determine that the item met the manufacturer's published specifications unless stated otherwise below.

Ambient Conditions: Temperature: 74.0 Degrees Fahrenheit Humidity: 49 % RH

Calibration Date: 17 Mar 99 **Calibration Procedure:** TO33K5-4-42-1 APR97

Condition as Received: IN TOLERANCE

Condition as Released: IN TOLERANCE

Remarks:

Approved by:

Jim Patterson, Supervisor or Walt Hill, Metrologist

Certificate # 33563

m:\a2la.rpt Rev date 10 Mar 99

Measurements performed by:

Mack Wood, Technician

Page 1 of 1



SOUTHWEST RESEARCH INSTITUTE™

6220 Culebra Road, P.O. Drawer 28510
Institute Quality Systems
Institute Calibration Laboratory
Phone: 210-522-5215 Fax 210-522-3692

Certificate of Calibration

Submitted By: DIV20

Address: B57

Contact: DARRELL DUNN

Manufacturer Model: FISHER SCIENTIFIC 15-166A

Description: THERMOMETER

Serial No: C98-132

Asset No: 007170

Procedure: CL-9, 5/99

Work Order: 444050127

Date Issued: Sep 17, 2002

Calibration Date: Sep 17, 2002

****Calibration Due:** Sep 17, 2003

Calibration Location: N/A

Environment: Temp. 74.0°F Hum. 50 %RH

***As Found:** IN TOLERANCE

***As Left:** IN TOLERANCE

This certificate documents traceability to the National Institute of Standards and Technology (NIST) and the International System of Units (SI). The Laboratory quality system conforms to ISO/IEC 17025, 1999 and ANSI/NCCL Z540-1-1994 which are equivalent to relevant requirements of the ISO 9000-1994 series of standards. This certificate may not be reproduced, except in full, without the written approval of the Southwest Research Institute Calibration Laboratory. The results of this calibration relate only to the individual instrument described above. This certificate shall not be used to claim product endorsement by the American Association for Laboratory Accreditation (A2LA) or any agency of the U. S. Government.

Uncertainty evaluation includes the item under test and is calculated in accordance with the ISO "Guide to the Expression of Uncertainty in Measurement" (GUM). The uncertainty represents an expanded uncertainty using a coverage factor of k=2 to approximate a 95% confidence level. The calibration process provides a Test Uncertainty Ratio (TUR) of less than or equal to 25% (4:1) of the test limit unless otherwise stated in remarks or an attachment.

*The client has sole responsibility for determination of in/out of tolerance or compliance/noncompliance. An in/out of tolerance opinion is provided for your convenience based only on the Test Instrument (TI) reading(s) and limits as reported. The reported uncertainty relates only to the results at the time of calibration and does not imply any short or long term stability of the TI.

**Calibration interval is determined by the client and does not assure the instrument will remain within tolerance until this date. Any number of factors may cause the instrument to be out of tolerance before the next calibration date.

Remarks: None

Standards Used

Asset	Manufacturer	Model	Description	Cal Due
009137	HART SCIENTIFIC, INC	1575	THERMOMETER	Dec 10, 02
008920	HART SCIENTIFIC, INC	17660-A-120-6-W	PLATINUM RTD	Dec 07, 02

Approved by: Walt Hill
Metrology Group Leader

Measurements by: Walt Hill
Metrology Technician

Southwest Research Institute
Calibration laboratory
Calibration Sheet.

Work Order:	444050127	Mfr.	FISHER SCIENTIFIC	Technician	WHILL
Asset No.	007170	Model	15-166A	Procedure	CL-9 5/99
Serial No.	C98-132	Type.	THERMOMETER	Cal Date.	September 17, 2002
Remarks:					

Function/Range	Standard	TI Reading	Difference	Test Limits+/-	Uncertainty	Found/Left
	Deg c	Deg c	Deg c	Deg c	Deg c	Results
-20	-19.81	-20.1	-0.3	1	1.2	Pass
0	0.11	0.0	-0.1	1	1.2	Pass
50	49.99	49.7	-0.3	1	1.2	Pass
100	99.94	100.0	0.1	1	1.2	Pass
150	149.83	149.0	-0.8	1	1.2	Pass



SOUTHWEST RESEARCH INSTITUTE™

6220 Culebra Road, P.O. Drawer 28510
Institute Quality Systems
Institute Calibration Laboratory
Phone: 210-522-5215 Fax 210-522-3692

Certificate of Calibration

Submitted By: DIV20

Address: B57

Contact: DARRELL DUNN

Manufacturer Model: FISHER SCIENTIFIC 15-166A

Description: THERMOMETER

Serial No: C98-132

Asset No: 007170

Procedure: TEMPERATURE, MAY/03

Work Order: 444055336

Date Issued: Sep 11, 2003

Calibration Date: Sep 10, 2003

****Calibration Due:** Sep 10, 2004

Calibration Location: Bldg. 64

Environment: Temp. 81.0°F Hum. 49 %RH

***As Found:** IN TOLERANCE

***As Left:** IN TOLERANCE

This certificate documents traceability to the National Institute of Standards and Technology (NIST) and the International System of Units (SI). The Laboratory quality system conforms to ISO/IEC 17025, 1999 and ANSI/NCCL Z540-1-1994 which are equivalent to relevant requirements of the ISO 9000-1994 series of standards. This certificate may not be reproduced, except in full, without the written approval of the Southwest Research Institute Calibration Laboratory. The results of this calibration relate only to the individual instrument described above. This certificate shall not be used to claim product endorsement by the American Association for Laboratory Accreditation (A2LA) or any agency of the U. S. Government.

Uncertainty evaluation includes the item under test and is calculated in accordance with the ISO "Guide to the Expression of Uncertainty in Measurement" (GUM). The uncertainty represents an expanded uncertainty using a coverage factor of k=2 to approximate a 95% confidence level. The calibration process provides a Test Uncertainty Ratio (TUR) of less than or equal to 25% (4:1) of the test limit unless otherwise stated in remarks or an attachment.

*The client has sole responsibility for determination of in/out of tolerance or compliance/noncompliance. An in/out of tolerance opinion is provided for your convenience based only on the Test Instrument (TI) reading(s) and limits as reported. The reported uncertainty relates only to the results at the time of calibration and does not imply any short or long term stability of the TI.

**Calibration interval is determined by the client and does not assure the instrument will remain within tolerance until this date. Any number of factors may cause the instrument to be out of tolerance before the next calibration date.

Remarks: None

Standards Used

Asset	Manufacturer	Model	Description	Cal Due
009137	HART SCIENTIFIC, INC	1575	THERMOMETER	Feb 05, 04
008920	HART SCIENTIFIC, INC	17660-A-120-6-W	PLATINUM RTD	Feb 07, 04

Approved by: Walt Hill
Metrology Group Leader
m:\Nona21a1.rpt Rev date 15, August 02

Measurements by: Mark Romero
Metrology Technician

Southwest Research Institute
 Calibration laboratory
 Calibration Report

Work Order:	444055336	Mfr.	Fisher Scientific	Technician	Mark Romero
Asset No.	007170	Model	15-166A	Cal Date.	10-Sep-03
Serial No.	C98-132	Type.	THERMOMETER		
Remarks:					

Function/Range	Test Point	TI Reading	Difference	+/-Limit	+/-Uncertainty	Found/Left
Temperature	Deg C	Deg C	Deg C	Deg C	Deg C	Result
	-19.9	-19.8	0.1	0.5	1.2	Pass
	0.1	0.5	0.4	0.5	1.2	Pass
	50.0	50.2	0.2	0.5	1.2	Pass
	100.0	100.2	0.2	0.5	1.2	Pass
	149.9	150.3	0.4	0.5	1.2	Pass
END OF REPORT						



SOUTHWEST RESEARCH INSTITUTE™

6220 Culebra Road, P.O. Drawer 28510
Institute Quality Systems
Institute Calibration Laboratory
Phone: 210-522-5215 Fax 210-522-3692



Certificate of Calibration

0972-01

Submitted By: DIV20

Address: B57

Contact: DARRELL DUNN

Manufacturer Model: FISHER SCIENTIFIC 15-166A

Description: THERMOMETER

Serial No: C98-132

Asset No: 007170

Procedure: TEMPERATURE, JAN/04

Work Order: 444060616

Date Issued: Sep 9, 2004

Calibration Date: Sep 9, 2004

****Calibration Due:** Sep 9, 2005

Calibration Location: Bldg. 64

Environment: Temp. 73.0°F Hum. 40 %RH

***As Found:** IN TOLERANCE

***As Left:** IN TOLERANCE

This certificate documents traceability to the National Institute of Standards and Technology (NIST) and the International System of Units (SI). The Laboratory quality system conforms to ISO/IEC 17025, 1999 and ANSI/NCCL Z540-1-1994 which are equivalent to relevant requirements of the ISO 9000-1994 series of standards. This certificate may not be reproduced, except in full, without the written approval of the Southwest Research Institute Calibration Laboratory. The results of this calibration relate only to the individual instrument described above. This certificate shall not be used to claim product endorsement by the American Association for Laboratory Accreditation (A2LA) or any agency of the U. S. Government.

Uncertainty evaluation includes the item under test and is calculated in accordance with the ISO "Guide to the Expression of Uncertainty in Measurement" (GUM). The uncertainty represents an expanded uncertainty using a coverage factor of k=2 to approximate a 95% confidence level. See Remarks or attached Calibration Report with the same Work Order number for calibration data.

*The client has sole responsibility for determination of in/out of tolerance or compliance/noncompliance. An in/out of tolerance opinion is provided for your convenience based only on the Test Instrument (TI) reading(s) and limits as reported. The reported uncertainty relates only to the results at the time of calibration and does not imply any short or long term stability of the TI.

**Calibration interval is determined by the client and does not assure the instrument will remain within tolerance until this date. Any number of factors may cause the instrument to be out of tolerance before the next calibration date.

Remarks: None

Standards Used

Asset	Manufacturer	Model	Description	Cal Due
009137	HART SCIENTIFIC	1575	THERMOMETER	Sep 25, 04
008920	HART SCIENTIFIC	5614-17660-A-12	PLATINUM RTD	Sep 25, 04

Approved by: Walt Hill
Metrology Group Leader
m:\a2la1.rpt Rev date 11, May 04

Measurements by: Bob Tromlinger
Metrology Technician

Southwest Research Institute
 Calibration Laboratory
 Measurement Report

Work Order:	444060616	Mfr.	Fisher-Scientific	Technician	blt
Asset No.	007170	Model	15-166A	Cal Date.	09-Sep-04
Serial No.	C98-132	Type.	THERMOMETER		
Remarks:	Accuracy Reference ANSI Z236.1-1983				

Function/Range	Test Point	TI Reading	Difference	+/-Limit	+/-Uncertainty	Found/Left
Temperature	Deg C	Deg C	Deg C	Deg C	Deg C	Result
-20	-19.9	-20.0	-0.1	0.5	0.12	Pass
0	0.1	0.0	-0.1	0.5	0.12	Pass
50	50.0	50.0	0.0	0.5	0.12	Pass
100	100.0	100.0	0.0	0.5	0.12	Pass
150	150.0	149.5	-0.5	0.5	0.12	Pass
END OF REPORT						