



Southwest Research Institute
6220 Culebra Road
San Antonio, TX 78238
(210) 522-5215
Department of Quality Assurance
Calibration Laboratory

Accredited

Certificate #
0972-01

Certificate of Calibration

17 March 1999

Issued to: DARRELL DUNN DIV20 B57
Manufacturer/Model: FISHER SCIENTIFIC 15-166A
Description: THERMOMETER
Serial Number: C98-101
Asset Number: 007168

This certifies the above item was calibrated in compliance with MIL-STD-45662A and ANSI/NC SL Z540-1-1994. Standards used in this calibration, described in the referenced calibration procedure with associated uncertainties or tolerances, are traceable to the National Institute of Standards and Technology (NIST). Supporting documentation relative to traceability is on file and is available for examination upon request. This certificate is not to be reproduced, except in full, without the written approval of the Southwest Research Institute Department of Quality Assurance Calibration Laboratory.

This laboratory is accredited by the American Association for Laboratory Accreditation (A2LA) and the results of this calibration certificate were determined in accordance with the terms of accreditation unless stated otherwise below.

The uncertainty of the calibration was sufficient to determine that the item met the manufacturer's published specifications unless stated otherwise below.

Ambient Conditions: Temperature: 74.0 Degrees Fahrenheit Humidity: 49 % RH


Calibration Date: 17 Mar 99 **Calibration Procedure:** TO33K5-4-42-1 APR97

Condition as Received: IN TOLERANCE

Condition as Released: IN TOLERANCE

Remarks:

Approved by:


Jim Patterson, Supervisor or Walt Hill, Metrologist

Certificate # 33565

m:\a2la.rpt Rev date 10 Mar 99

Measurements performed by:


Mack Wood, Technician

Page 1 of 1