



Southwest Research Institute
6220 Culebra Road
San Antonio, TX 78238
(210) 522-5215
Department of Quality Assurance
Calibration Laboratory



Certificate #
0972-01

Certificate of Calibration

17 March 1999

Issued to: DARRELL DUNN DIV20 B57
Manufacturer/Model: FISHER SCIENTIFIC 15-166A
Description: THERMOMETER
Serial Number: H98-179
Asset Number: 007169

This certifies the above item was calibrated in compliance with MIL-STD-45662A and ANSI/NCSL Z540-1-1994. Standards used in this calibration, described in the referenced calibration procedure with associated uncertainties or tolerances, are traceable to the National Institute of Standards and Technology (NIST). Supporting documentation relative to traceability is on file and is available for examination upon request. This certificate is not to be reproduced, except in full, without the written approval of the Southwest Research Institute Department of Quality Assurance Calibration Laboratory.

This laboratory is accredited by the American Association for Laboratory Accreditation (A2LA) and the results of this calibration certificate were determined in accordance with the terms of accreditation unless stated otherwise below.

The uncertainty of the calibration was sufficient to determine that the item met the manufacturer's published specifications unless stated otherwise below.

Ambient Conditions: Temperature: 74.0 Degrees Fahrenheit Humidity: 49 % RH


Calibration Date: 17 Mar 99 **Calibration Procedure:** TO33K5-4-42-1 APR97

Condition as Received: IN TOLERANCE

Condition as Released: IN TOLERANCE

Remarks:


Approved by:


Jim Patterson, Supervisor or Walt Hill, Metrologist

Certificate # 33564

m:\a2la.rpt Rev date 10 Mar 99

Measurements performed by:


Mack Wood, Technician

Page 1 of 1

SOUTHWEST RESEARCH INSTITUTE

6220 CULEBRA ROAD • POST OFFICE DRAWER 28510 • SAN ANTONIO, TEXAS, 78228-0510 • TEL (210) 522-5215 • FAX (210) 522-3692

To: Darrel Dunn, Div 20, Bldg 57
From: Walt Hill, Institute Calibration Laboratory Supervisor
CC: Rodney Weber, Institute Quality Assurance Manager
Date: May. 1, 01
Subject: Out-of-tolerance Notice

The purpose of this notice is to alert you of a condition, which could have caused erroneous measurements. The as-found readings are provided for your evaluation to determine if the instrument had any impact on your operations and if further action is required. If we can be of assistance, please contact the Calibration Laboratory at 522-5215.

Manufacturer: Fischer Scientific **Model:** 15-166A

Description: Thermometer **Serial Number:** H98-179

Asset Number: 7169 **User ID Number:**

Last Calibration: 4/18/00

Date Received for Service: Apr. 26, 01 **Work Order Number:** 444043350

Service Requested: Calibrate before use

Remarks: Thermometer has partial separation of mercury . As Found taken in this condition.

AS-FOUND DATA

PARAMETER OR FUNCTION	APPLIED OR NOMINAL VALUE	INSTRUMENT READING	INSTRUMENT ERROR	INSTRUMENT TOLERANCE
Temperature	-19.89 Deg C	-20.6 Deg C	0.74 Deg C	+/- 1 Deg C
	0.000 Deg C	-0.3 Deg C	0.3 Deg C	+/- 1 Deg C
	50.000 Deg C	50.5 Deg C	0.5 Deg C	+/- 1 Deg C
	100.000 Deg C	99.0 Deg C	1.0 Deg C	+/- 1 Deg C
	149.966 Deg C	148.6 Deg C	1.36 Deg C	+/-1 Deg C

OUT OF TOLERANCE



SOUTHWEST RESEARCH INSTITUTE™

6220 Culebra Road, P.O. Drawer 28510
Institute Quality Systems
Institute Calibration Laboratory
Phone: 210-522-5215 Fax 210-522-3692

Certificate of Calibration

Submitted By: DIV20

Address: B57

Contact: DARRELL DUNN

Manufacturer Model: FISHER SCIENTIFIC 15-166A

Description: THERMOMETER

Serial No: H98-179

Asset No: 007169

Procedure: THERMOMETERS, JAN/03

Work Order: 444053497

Date Issued: Apr 29, 2003

Calibration Date: Apr 29, 2003

****Calibration Due:** Apr 29, 2004

Calibration Location: Bldg. 64

Environment: Temp. 78.0°F Hum. 47 %RH

***As Found:** IN TOLERANCE

***As Left:** IN TOLERANCE

This certificate documents traceability to the National Institute of Standards and Technology (NIST) and the International System of Units (SI). The Laboratory quality system conforms to ISO/IEC 17025, 1999 and ANSI/NCSL Z540-1-1994 which are equivalent to relevant requirements of the ISO 9000-1994 series of standards. This certificate may not be reproduced, except in full, without the written approval of the Southwest Research Institute Calibration Laboratory. The results of this calibration relate only to the individual instrument described above. This certificate shall not be used to claim product endorsement by the American Association for Laboratory Accreditation (A2LA) or any agency of the U. S. Government.

Uncertainty evaluation includes the item under test and is calculated in accordance with the ISO "Guide to the Expression of Uncertainty in Measurement" (GUM). The uncertainty represents an expanded uncertainty using a coverage factor of $k=2$ to approximate a 95% confidence level. The calibration process provides a Test Uncertainty Ratio (TUR) of less than or equal to 25% (4:1) of the test limit unless otherwise stated in remarks or an attachment.


*The client has sole responsibility for determination of in/out of tolerance or compliance/noncompliance. An in/out of tolerance opinion is provided for your convenience based only on the Test Instrument (TI) reading(s) and limits as reported. The reported uncertainty relates only to the results at the time of calibration and does not imply any short or long term stability of the TI.


**Calibration interval is determined by the client and does not assure the instrument will remain within tolerance until this date. Any number of factors may cause the instrument to be out of tolerance before the next calibration date.

Remarks: None

Standards Used

Asset	Manufacturer	Model	Description	Cal Due
009137	HART SCIENTIFIC, INC	1575	THERMOMETER	Jul 06, 03
008920	HART SCIENTIFIC, INC	17660-A-120-6-W	PLATINUM RTD	Jul 06, 03

Approved by: 
Metrology Group Leader
m:\Nona2\al.rpt Rev date 15, August 02


Measurements by: Mark Romero
Metrology Technician

Southwest Research Institute
Calibration laboratory
Measurement Record

Work Order:	444053497	Mfr.	Fisher Scientific	Technician	Mark Romero
Asset No.	007169	Model	15-166A		
Serial No.	H98-179	Type.	THERMOMETER	Cal Date.	29-Apr-03
Remarks:					

Function/Range	Test Point	TI Reading	Difference	+/-Limits	+/-Uncertainty	Found/Left
Temperature	Deg C	Deg C	Deg C	Deg C	Deg C	Results
-20	-19.9	-20.1	-0.2	0.5	0.6	Pass
0	0.1	0.0	-0.1	0.5	0.6	Pass
50	50.0	49.7	-0.3	0.5	0.6	Pass
100	100.0	99.6	-0.4	0.5	0.6	Pass
150	149.9	149.7	-0.2	0.5	0.6	Pass
END OF REPORT						



SOUTHWEST RESEARCH INSTITUTE™

6220 Culebra Road, P.O. Drawer 28510
Institute Quality Systems
Institute Calibration Laboratory
Phone: 210-522-5215 Fax 210-522-3692



0972-01

Certificate of Calibration

Submitted By: DIV20

Address: B57

Contact: DARRELL DUNN

Manufacturer Model: FISHER SCIENTIFIC 15-166A

Description: THERMOMETER

Serial No: H98-179

Asset No: 007169

Procedure: TEMPERATURE, JAN/04

Work Order: 444058713

Date Issued: Apr 28, 2004

Calibration Date: Apr 28, 2004

****Calibration Due:** Apr 28, 2005

Calibration Location: Bldg. 64

Environment: Temp. 80.0°F Hum. 38 %RH

***As Found:** IN TOLERANCE

***As Left:** IN TOLERANCE

This certificate documents traceability to the National Institute of Standards and Technology (NIST) and the International System of Units (SI). The Laboratory quality system conforms to ISO/IEC 17025, 1999 and ANSI/NCCL Z540-1-1994 which are equivalent to relevant requirements of the ISO 9000-1994 series of standards. This certificate may not be reproduced, except in full, without the written approval of the Southwest Research Institute Calibration Laboratory. The results of this calibration relate only to the individual instrument described above. This certificate shall not be used to claim product endorsement by the American Association for Laboratory Accreditation (A2LA) or any agency of the U. S. Government.

Uncertainty evaluation includes the item under test and is calculated in accordance with the ISO "Guide to the Expression of Uncertainty in Measurement" (GUM). The uncertainty represents an expanded uncertainty using a coverage factor of $k=2$ to approximate a 95% confidence level. The calibration process provides a Test Uncertainty Ratio (TUR) of less than or equal to 25% (4:1) of the test limit unless otherwise stated in remarks or an attachment.

*The client has sole responsibility for determination of in/out of tolerance or compliance/noncompliance. An in/out of tolerance opinion is provided for your convenience based only on the Test Instrument (TI) reading(s) and limits as reported. The reported uncertainty relates only to the results at the time of calibration and does not imply any short or long term stability of the TI.

**Calibration interval is determined by the client and does not assure the instrument will remain within tolerance until this date. Any number of factors may cause the instrument to be out of tolerance before the next calibration date.

Remarks: None

Standards Used

Asset	Manufacturer	Model	Description	Cal Due
009137	HART SCIENTIFIC	1575	THERMOMETER	Sep 05, 04

Approved by: Walt Hill
Metrology Group Leader
m:\a2la1.rpt Rev date 15, August 02

Measurements by: Scott Kester
Metrology Technician

Southwest Research Institute
Calibration laboratory
Caibration Report

Work Order:	444058713	Mfr.	Fisher-Scientific	Technician	SRK
Asset No.	007169	Model	15-166A		
Serial No.	H98-179	Type.	THERMOMETER	Cal Date.	28-Apr-04
Remarks:					

Function/Range	Test Point	TI Reading	Difference	+/-Limit	+/-Uncertainty	Found/Left
Temperature	Deg C	Deg C	Deg C	Deg C	Deg C	Result
-20	-20.0	-20.0	0.0	0.5	0.58	Pass
0	0.1	0.5	0.4	0.5	0.58	Pass
50	50.0	49.8	-0.3	0.5	0.58	Pass
100	99.9	99.5	-0.4	0.5	0.58	Pass
150	149.9	149.5	-0.4	0.5	0.58	Pass
END OF REPORT						