

S O U T H W E S T R E S E A R C H I N S T I T U T E

Department of Quality Assurance
Calibration Laboratory

Device Serial No: 115820

Calibration Date: 07/22/94

STANDARDS

Standard No: 219 Manufacturer: AZONIX

Nomenclature: RTD THERMOMETER

Serial No: T1587-2078

Cal.Due: 02/22/95

Model: A1011

Cal.Rec.No: 00013377

Standard No: 328 Manufacturer: AZONIX

Nomenclature: RTD TEMPERATURE PROBE

Serial No: 351477

Cal.Due: 03/10/95

Model: A12001

Cal.Rec.No: 00013378

SOUTHWEST RESEARCH INSTITUTE
Department of Quality Assurance
Calibration Laboratory

CERTIFICATE OF CALIBRATION
07/28/95

Issued to: DARRELL DUNN DIV20 ,B57
Manufacturer: KESSL
Nomenclature: THERMOMETER
Serial Number: 115820
Notes:

Asset Number: 003238
Model Number: ASTM IC 76MM

SwRI/Div. I.D. #: NONE

ENVIRONMENTAL CONDITIONS

Temperature: 74.0F

Relative Humidity: 60%

CALIBRATION INFORMATION

Procedure Number: WI-9-30-TH03
Remarks:

Accuracy: MFGR SPECS
Received IN Tolerance

**Calibration was in accordance with requirements of MIL-STD-45662A.
Measurements are traceable to the National Institute of Standards and Tech-
nology. Inspection and test data are on file and available for inspection.**

STANDARDS USED FOR CERTIFICATION

Asset #	Serial #	Mfg	Model #	Nomenclature	Cal Date	Int.	Cal Due
000219	T1587-2078	AZONI	A1011	RTD THERMOMETER	03/06/95	12	03/06/96
000328	351477	AZONI	A12001	RTD TEMPERATURE PROBE	03/06/95	12	03/06/96

Certified by :



Certificate#: 18122

Calibration Date: 07/28/95
Interval: 12 months
Next Calibration Due: 07/28/96

